

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3673

Continuing Calib Date: 11/22/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 03:43

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.48	5.49	5.39	5.59	0.01
Aroclor-1248-2 (2)	5.75	5.76	5.66	5.86	0.01
Aroclor-1248-3 (3)	5.95	5.96	5.86	6.06	0.01
Aroclor-1248-4 (4)	6.35	6.35	6.25	6.45	0.00
Aroclor-1248-5 (5)	6.38	6.39	6.29	6.49	0.01
Tetrachloro-m-xylene	4.34	4.35	4.25	4.45	0.01
Decachlorobiphenyl	9.92	9.93	9.83	10.03	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3673

Continuing Calib Date: 11/22/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 03:43

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1248-2 (2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-3 (3)	5.00	5.01	4.91	5.11	0.01
Aroclor-1248-4 (4)	5.17	5.18	5.08	5.28	0.01
Aroclor-1248-5 (5)	5.56	5.57	5.47	5.67	0.01
Tetrachloro-m-xylene	3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01