

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115255.D **Time Analyzed:** 10:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	540.560	500.000	8.1
Aroclor-1016-2	5.501	5.408	5.608	536.050	500.000	7.2
Aroclor-1016-3	5.562	5.470	5.670	552.900	500.000	10.6
Aroclor-1016-4	5.657	5.566	5.766	559.360	500.000	11.9
Aroclor-1016-5	5.949	5.858	6.058	521.290	500.000	4.3
Aroclor-1260-1	7.063	6.971	7.171	517.640	500.000	3.5
Aroclor-1260-2	7.316	7.225	7.425	562.740	500.000	12.5
Aroclor-1260-3	7.674	7.582	7.782	559.690	500.000	11.9
Aroclor-1260-4	7.895	7.803	8.003	542.460	500.000	8.5
Aroclor-1260-5	8.202	8.110	8.310	549.580	500.000	9.9
Decachlorobiphenyl	9.917	9.830	10.030	55.630	50.000	11.3
Tetrachloro-m-xylene	4.340	4.248	4.448	54.910	50.000	9.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115255.D **Time Analyzed:** 10:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.628	4.828	556.230	500.000	11.2
Aroclor-1016-2	4.738	4.647	4.847	510.610	500.000	2.1
Aroclor-1016-3	4.913	4.821	5.021	527.010	500.000	5.4
Aroclor-1016-4	4.955	4.864	5.064	521.420	500.000	4.3
Aroclor-1016-5	5.167	5.076	5.276	533.170	500.000	6.6
Aroclor-1260-1	6.196	6.106	6.306	490.360	500.000	-1.9
Aroclor-1260-2	6.384	6.294	6.494	476.010	500.000	-4.8
Aroclor-1260-3	6.536	6.446	6.646	489.220	500.000	-2.2
Aroclor-1260-4	7.007	6.917	7.117	500.090	500.000	0.0
Aroclor-1260-5	7.249	7.158	7.358	508.880	500.000	1.8
Decachlorobiphenyl	8.629	8.541	8.741	56.900	50.000	13.8
Tetrachloro-m-xylene	3.643	3.548	3.748	53.730	50.000	7.5