

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115271.D **Time Analyzed:** 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.480	5.387	5.587	511.280	500.000	2.3
Aroclor-1016-2	5.502	5.408	5.608	505.380	500.000	1.1
Aroclor-1016-3	5.563	5.470	5.670	528.620	500.000	5.7
Aroclor-1016-4	5.659	5.566	5.766	516.660	500.000	3.3
Aroclor-1016-5	5.950	5.858	6.058	501.060	500.000	0.2
Aroclor-1260-1	7.064	6.971	7.171	492.400	500.000	-1.5
Aroclor-1260-2	7.317	7.225	7.425	529.780	500.000	6.0
Aroclor-1260-3	7.674	7.582	7.782	529.030	500.000	5.8
Aroclor-1260-4	7.895	7.803	8.003	518.210	500.000	3.6
Aroclor-1260-5	8.202	8.110	8.310	514.380	500.000	2.9
Decachlorobiphenyl	9.917	9.830	10.030	52.250	50.000	4.5
Tetrachloro-m-xylene	4.341	4.248	4.448	51.870	50.000	3.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115271.D **Time Analyzed:** 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.628	4.828	513.940	500.000	2.8
Aroclor-1016-2	4.740	4.647	4.847	487.770	500.000	-2.4
Aroclor-1016-3	4.914	4.821	5.021	496.740	500.000	-0.7
Aroclor-1016-4	4.957	4.864	5.064	493.770	500.000	-1.2
Aroclor-1016-5	5.168	5.076	5.276	501.380	500.000	0.3
Aroclor-1260-1	6.198	6.106	6.306	459.360	500.000	-8.1
Aroclor-1260-2	6.386	6.294	6.494	444.420	500.000	-11.1
Aroclor-1260-3	6.538	6.446	6.646	455.360	500.000	-8.9
Aroclor-1260-4	7.009	6.917	7.117	464.150	500.000	-7.2
Aroclor-1260-5	7.250	7.158	7.358	469.900	500.000	-6.0
Decachlorobiphenyl	8.631	8.541	8.741	52.920	50.000	5.8
Tetrachloro-m-xylene	3.645	3.548	3.748	50.070	50.000	0.1