

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115287.D **Time Analyzed:** 23:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.478	5.387	5.587	519.920	500.000	4.0
Aroclor-1016-2	5.500	5.408	5.608	508.370	500.000	1.7
Aroclor-1016-3	5.561	5.470	5.670	532.720	500.000	6.5
Aroclor-1016-4	5.657	5.566	5.766	519.080	500.000	3.8
Aroclor-1016-5	5.948	5.858	6.058	510.590	500.000	2.1
Aroclor-1260-1	7.062	6.971	7.171	511.360	500.000	2.3
Aroclor-1260-2	7.315	7.225	7.425	542.750	500.000	8.6
Aroclor-1260-3	7.672	7.582	7.782	537.370	500.000	7.5
Aroclor-1260-4	7.895	7.803	8.003	523.740	500.000	4.7
Aroclor-1260-5	8.200	8.110	8.310	524.150	500.000	4.8
Decachlorobiphenyl	9.916	9.830	10.030	54.210	50.000	8.4
Tetrachloro-m-xylene	4.340	4.248	4.448	51.960	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115287.D **Time Analyzed:** 23:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.628	4.828	492.720	500.000	-1.5
Aroclor-1016-2	4.739	4.647	4.847	480.960	500.000	-3.8
Aroclor-1016-3	4.912	4.821	5.021	500.290	500.000	0.1
Aroclor-1016-4	4.955	4.864	5.064	491.120	500.000	-1.8
Aroclor-1016-5	5.167	5.076	5.276	528.530	500.000	5.7
Aroclor-1260-1	6.196	6.106	6.306	464.970	500.000	-7.0
Aroclor-1260-2	6.383	6.294	6.494	455.540	500.000	-8.9
Aroclor-1260-3	6.536	6.446	6.646	464.390	500.000	-7.1
Aroclor-1260-4	7.007	6.917	7.117	469.760	500.000	-6.0
Aroclor-1260-5	7.248	7.158	7.358	474.240	500.000	-5.2
Decachlorobiphenyl	8.629	8.541	8.741	53.290	50.000	6.6
Tetrachloro-m-xylene	3.644	3.548	3.748	50.640	50.000	1.3