

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115302.D **Time Analyzed:** 05:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	527.260	500.000	5.5
Aroclor-1016-2	5.500	5.408	5.608	521.770	500.000	4.4
Aroclor-1016-3	5.562	5.470	5.670	551.380	500.000	10.3
Aroclor-1016-4	5.658	5.566	5.766	525.150	500.000	5.0
Aroclor-1016-5	5.949	5.858	6.058	516.580	500.000	3.3
Aroclor-1260-1	7.062	6.971	7.171	521.870	500.000	4.4
Aroclor-1260-2	7.316	7.225	7.425	548.770	500.000	9.8
Aroclor-1260-3	7.673	7.582	7.782	536.500	500.000	7.3
Aroclor-1260-4	7.895	7.803	8.003	522.080	500.000	4.4
Aroclor-1260-5	8.201	8.110	8.310	524.670	500.000	4.9
Decachlorobiphenyl	9.916	9.830	10.030	54.610	50.000	9.2
Tetrachloro-m-xylene	4.340	4.248	4.448	53.020	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115302.D **Time Analyzed:** 05:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.628	4.828	494.910	500.000	-1.0
Aroclor-1016-2	4.739	4.647	4.847	516.910	500.000	3.4
Aroclor-1016-3	4.913	4.821	5.021	508.040	500.000	1.6
Aroclor-1016-4	4.955	4.864	5.064	492.570	500.000	-1.5
Aroclor-1016-5	5.167	5.076	5.276	548.710	500.000	9.7
Aroclor-1260-1	6.197	6.106	6.306	476.950	500.000	-4.6
Aroclor-1260-2	6.384	6.294	6.494	457.320	500.000	-8.5
Aroclor-1260-3	6.537	6.446	6.646	455.490	500.000	-8.9
Aroclor-1260-4	7.007	6.917	7.117	463.830	500.000	-7.2
Aroclor-1260-5	7.249	7.158	7.358	474.670	500.000	-5.1
Decachlorobiphenyl	8.630	8.541	8.741	53.030	50.000	6.1
Tetrachloro-m-xylene	3.644	3.548	3.748	51.250	50.000	2.5