

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115309.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.478	5.387	5.587	535.940	500.000	7.2
Aroclor-1016-2	5.500	5.408	5.608	527.310	500.000	5.5
Aroclor-1016-3	5.561	5.470	5.670	556.270	500.000	11.3
Aroclor-1016-4	5.657	5.566	5.766	537.270	500.000	7.5
Aroclor-1016-5	5.949	5.858	6.058	527.120	500.000	5.4
Aroclor-1260-1	7.062	6.971	7.171	540.080	500.000	8.0
Aroclor-1260-2	7.316	7.225	7.425	568.980	500.000	13.8
Aroclor-1260-3	7.673	7.582	7.782	558.180	500.000	11.6
Aroclor-1260-4	7.894	7.803	8.003	539.790	500.000	8.0
Aroclor-1260-5	8.201	8.110	8.310	539.600	500.000	7.9
Decachlorobiphenyl	9.916	9.830	10.030	54.540	50.000	9.1
Tetrachloro-m-xylene	4.340	4.248	4.448	53.170	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115309.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.628	4.828	515.750	500.000	3.2
Aroclor-1016-2	4.739	4.647	4.847	515.840	500.000	3.2
Aroclor-1016-3	4.913	4.821	5.021	516.350	500.000	3.3
Aroclor-1016-4	4.955	4.864	5.064	498.640	500.000	-0.3
Aroclor-1016-5	5.167	5.076	5.276	561.060	500.000	12.2
Aroclor-1260-1	6.196	6.106	6.306	489.880	500.000	-2.0
Aroclor-1260-2	6.384	6.294	6.494	478.550	500.000	-4.3
Aroclor-1260-3	6.536	6.446	6.646	479.370	500.000	-4.1
Aroclor-1260-4	7.007	6.917	7.117	475.700	500.000	-4.9
Aroclor-1260-5	7.248	7.158	7.358	485.190	500.000	-3.0
Decachlorobiphenyl	8.629	8.541	8.741	54.030	50.000	8.1
Tetrachloro-m-xylene	3.644	3.548	3.748	51.660	50.000	3.3