

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115324.D **Time Analyzed:** 14:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.480	5.387	5.587	526.860	500.000	5.4
Aroclor-1016-2	5.502	5.408	5.608	520.910	500.000	4.2
Aroclor-1016-3	5.563	5.470	5.670	539.130	500.000	7.8
Aroclor-1016-4	5.658	5.566	5.766	516.940	500.000	3.4
Aroclor-1016-5	5.950	5.858	6.058	528.680	500.000	5.7
Aroclor-1260-1	7.063	6.971	7.171	518.970	500.000	3.8
Aroclor-1260-2	7.318	7.225	7.425	554.590	500.000	10.9
Aroclor-1260-3	7.674	7.582	7.782	537.980	500.000	7.6
Aroclor-1260-4	7.896	7.803	8.003	519.110	500.000	3.8
Aroclor-1260-5	8.202	8.110	8.310	523.890	500.000	4.8
Decachlorobiphenyl	9.920	9.830	10.030	54.080	50.000	8.2
Tetrachloro-m-xylene	4.341	4.248	4.448	54.550	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115324.D **Time Analyzed:** 14:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.628	4.828	523.850	500.000	4.8
Aroclor-1016-2	4.739	4.647	4.847	523.880	500.000	4.8
Aroclor-1016-3	4.914	4.821	5.021	524.800	500.000	5.0
Aroclor-1016-4	4.955	4.864	5.064	507.740	500.000	1.5
Aroclor-1016-5	5.168	5.076	5.276	564.920	500.000	13.0
Aroclor-1260-1	6.197	6.106	6.306	481.310	500.000	-3.7
Aroclor-1260-2	6.385	6.294	6.494	461.930	500.000	-7.6
Aroclor-1260-3	6.538	6.446	6.646	465.230	500.000	-7.0
Aroclor-1260-4	7.008	6.917	7.117	453.180	500.000	-9.4
Aroclor-1260-5	7.249	7.158	7.358	457.000	500.000	-8.6
Decachlorobiphenyl	8.630	8.541	8.741	50.650	50.000	1.3
Tetrachloro-m-xylene	3.644	3.548	3.748	53.170	50.000	6.3