

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL09 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071364.D **Time Analyzed:** 11:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.497	4.397	4.597	526.880	500.000	5.4
Aroclor-1016-2	4.516	4.416	4.616	537.410	500.000	7.5
Aroclor-1016-3	4.574	4.474	4.674	526.150	500.000	5.2
Aroclor-1016-4	4.664	4.564	4.764	531.960	500.000	6.4
Aroclor-1016-5	4.950	4.849	5.049	539.290	500.000	7.9
Aroclor-1260-1	6.042	5.943	6.143	503.460	500.000	0.7
Aroclor-1260-2	6.301	6.200	6.400	536.860	500.000	7.4
Aroclor-1260-3	6.653	6.552	6.752	525.830	500.000	5.2
Aroclor-1260-4	6.867	6.767	6.967	528.400	500.000	5.7
Aroclor-1260-5	7.175	7.075	7.275	524.110	500.000	4.8
Decachlorobiphenyl	8.537	8.437	8.637	50.170	50.000	0.3
Tetrachloro-m-xylene	3.412	3.311	3.511	54.640	50.000	9.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL09 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071364.D **Time Analyzed:** 11:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.870	3.771	3.971	580.420	500.000	16.1
Aroclor-1016-2	3.887	3.788	3.988	585.740	500.000	17.1
Aroclor-1016-3	4.048	3.950	4.150	576.350	500.000	15.3
Aroclor-1016-4	4.097	3.998	4.198	569.100	500.000	13.8
Aroclor-1016-5	4.294	4.195	4.395	580.470	500.000	16.1
Aroclor-1260-1	5.295	5.197	5.397	584.340	500.000	16.9
Aroclor-1260-2	5.484	5.386	5.586	586.530	500.000	17.3
Aroclor-1260-3	5.627	5.530	5.730	578.190	500.000	15.6
Aroclor-1260-4	6.091	5.993	6.193	581.360	500.000	16.3
Aroclor-1260-5	6.336	6.238	6.438	577.820	500.000	15.6
Decachlorobiphenyl	7.656	7.559	7.759	55.670	50.000	11.3
Tetrachloro-m-xylene	2.856	2.756	2.956	58.620	50.000	17.2