

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL10 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071375.D **Time Analyzed:** 16:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.502	4.397	4.597	489.780	500.000	-2.0
Aroclor-1016-2	4.521	4.416	4.616	493.370	500.000	-1.3
Aroclor-1016-3	4.578	4.474	4.674	486.060	500.000	-2.8
Aroclor-1016-4	4.669	4.564	4.764	491.670	500.000	-1.7
Aroclor-1016-5	4.954	4.849	5.049	497.090	500.000	-0.6
Aroclor-1260-1	6.048	5.943	6.143	471.800	500.000	-5.6
Aroclor-1260-2	6.304	6.200	6.400	498.960	500.000	-0.2
Aroclor-1260-3	6.657	6.552	6.752	489.190	500.000	-2.2
Aroclor-1260-4	6.871	6.767	6.967	490.430	500.000	-1.9
Aroclor-1260-5	7.179	7.075	7.275	494.910	500.000	-1.0
Decachlorobiphenyl	8.542	8.437	8.637	43.910	50.000	-12.2
Tetrachloro-m-xylene	3.416	3.311	3.511	47.300	50.000	-5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3673
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL10 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071375.D **Time Analyzed:** 16:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.869	3.771	3.971	549.230	500.000	9.8
Aroclor-1016-2	3.886	3.788	3.988	559.440	500.000	11.9
Aroclor-1016-3	4.048	3.950	4.150	548.730	500.000	9.7
Aroclor-1016-4	4.096	3.998	4.198	539.800	500.000	8.0
Aroclor-1016-5	4.294	4.195	4.395	554.830	500.000	11.0
Aroclor-1260-1	5.295	5.197	5.397	571.510	500.000	14.3
Aroclor-1260-2	5.485	5.386	5.586	589.950	500.000	18.0
Aroclor-1260-3	5.628	5.530	5.730	587.750	500.000	17.6
Aroclor-1260-4	6.091	5.993	6.193	572.290	500.000	14.5
Aroclor-1260-5	6.336	6.238	6.438	557.630	500.000	11.5
Decachlorobiphenyl	7.657	7.559	7.759	50.970	50.000	1.9
Tetrachloro-m-xylene	2.854	2.756	2.956	52.670	50.000	5.3