

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3674

Continuing Calib Date: 11/22/2025

Initial Calibration Date(s): 11/18/2025

11/18/2025

Continuing Calib Time: 01:53

Initial Calibration Time(s): 12:13

17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-2	(2)	4.52	4.52	4.42	4.62	0.01
Aroclor-1016-3	(3)	4.57	4.57	4.47	4.67	0.00
Aroclor-1016-4	(4)	4.66	4.66	4.56	4.76	0.00
Aroclor-1016-5	(5)	4.95	4.95	4.85	5.05	0.00
Aroclor-1260-1	(1)	6.04	6.04	5.94	6.14	0.00
Aroclor-1260-2	(2)	6.30	6.30	6.20	6.40	0.00
Aroclor-1260-3	(3)	6.65	6.65	6.55	6.75	0.00
Aroclor-1260-4	(4)	6.87	6.87	6.77	6.97	0.00
Aroclor-1260-5	(5)	7.18	7.18	7.08	7.28	0.00
Tetrachloro-m-xylene		3.41	3.41	3.31	3.51	0.00
Decachlorobiphenyl		8.53	8.54	8.44	8.64	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3674

Continuing Calib Date: 11/22/2025

Initial Calibration Date(s): 11/18/2025

11/18/2025

Continuing Calib Time: 01:53

Initial Calibration Time(s): 12:13

17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.87	3.87	3.77	3.97	0.00
Aroclor-1016-2	(2)	3.88	3.89	3.79	3.99	0.01
Aroclor-1016-3	(3)	4.05	4.05	3.95	4.15	0.00
Aroclor-1016-4	(4)	4.09	4.10	4.00	4.20	0.01
Aroclor-1016-5	(5)	4.29	4.30	4.20	4.40	0.01
Aroclor-1260-1	(1)	5.29	5.30	5.20	5.40	0.01
Aroclor-1260-2	(2)	5.48	5.49	5.39	5.59	0.01
Aroclor-1260-3	(3)	5.62	5.63	5.53	5.73	0.01
Aroclor-1260-4	(4)	6.09	6.09	5.99	6.19	0.00
Aroclor-1260-5	(5)	6.33	6.34	6.24	6.44	0.01
Tetrachloro-m-xylene		2.85	2.86	2.76	2.96	0.01
Decachlorobiphenyl		7.65	7.66	7.56	7.76	0.01