

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** TETR16  
**Lab Code:** ACE **SDG NO.:** Q3674  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 11/21/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PQ071418.D **Time Analyzed:** 09:06

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.511 | 4.397     | 4.597 | 468.710            | 500.000           | -6.3  |
| Aroclor-1016-2       | 4.530 | 4.416     | 4.616 | 456.980            | 500.000           | -8.6  |
| Aroclor-1016-3       | 4.588 | 4.474     | 4.674 | 480.010            | 500.000           | -4.0  |
| Aroclor-1016-4       | 4.678 | 4.564     | 4.764 | 465.540            | 500.000           | -6.9  |
| Aroclor-1016-5       | 4.963 | 4.849     | 5.049 | 473.390            | 500.000           | -5.3  |
| Aroclor-1260-1       | 6.057 | 5.943     | 6.143 | 463.600            | 500.000           | -7.3  |
| Aroclor-1260-2       | 6.315 | 6.200     | 6.400 | 479.950            | 500.000           | -4.0  |
| Aroclor-1260-3       | 6.667 | 6.552     | 6.752 | 470.240            | 500.000           | -6.0  |
| Aroclor-1260-4       | 6.882 | 6.767     | 6.967 | 482.260            | 500.000           | -3.5  |
| Aroclor-1260-5       | 7.189 | 7.075     | 7.275 | 484.030            | 500.000           | -3.2  |
| Decachlorobiphenyl   | 8.554 | 8.437     | 8.637 | 44.460             | 50.000            | -11.1 |
| Tetrachloro-m-xylene | 3.424 | 3.311     | 3.511 | 46.350             | 50.000            | -7.3  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** TETR16  
**Lab Code:** ACE **SDG NO.:** Q3674  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 11/21/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PQ071418.D **Time Analyzed:** 09:06

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 3.868 | 3.771                | 3.971 | 569.930            | 500.000           | 14.0 |
| Aroclor-1016-2       | 3.885 | 3.788                | 3.988 | 578.700            | 500.000           | 15.7 |
| Aroclor-1016-3       | 4.046 | 3.950                | 4.150 | 567.410            | 500.000           | 13.5 |
| Aroclor-1016-4       | 4.094 | 3.998                | 4.198 | 540.350            | 500.000           | 8.1  |
| Aroclor-1016-5       | 4.292 | 4.195                | 4.395 | 557.710            | 500.000           | 11.5 |
| Aroclor-1260-1       | 5.293 | 5.197                | 5.397 | 547.750            | 500.000           | 9.6  |
| Aroclor-1260-2       | 5.483 | 5.386                | 5.586 | 553.130            | 500.000           | 10.6 |
| Aroclor-1260-3       | 5.626 | 5.530                | 5.730 | 541.900            | 500.000           | 8.4  |
| Aroclor-1260-4       | 6.091 | 5.993                | 6.193 | 583.940            | 500.000           | 16.8 |
| Aroclor-1260-5       | 6.336 | 6.238                | 6.438 | 565.980            | 500.000           | 13.2 |
| Decachlorobiphenyl   | 7.659 | 7.559                | 7.759 | 53.390             | 50.000            | 6.8  |
| Tetrachloro-m-xylene | 2.854 | 2.756                | 2.956 | 53.020             | 50.000            | 6.0  |