

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071442.D **Time Analyzed:** 17:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	481.640	500.000	-3.7
Aroclor-1016-2	4.515	4.416	4.616	483.050	500.000	-3.4
Aroclor-1016-3	4.573	4.474	4.674	471.010	500.000	-5.8
Aroclor-1016-4	4.663	4.564	4.764	472.740	500.000	-5.5
Aroclor-1016-5	4.949	4.849	5.049	480.980	500.000	-3.8
Aroclor-1260-1	6.043	5.943	6.143	483.480	500.000	-3.3
Aroclor-1260-2	6.299	6.200	6.400	498.980	500.000	-0.2
Aroclor-1260-3	6.652	6.552	6.752	479.010	500.000	-4.2
Aroclor-1260-4	6.867	6.767	6.967	481.290	500.000	-3.7
Aroclor-1260-5	7.175	7.075	7.275	486.170	500.000	-2.8
Decachlorobiphenyl	8.537	8.437	8.637	43.880	50.000	-12.2
Tetrachloro-m-xylene	3.410	3.311	3.511	46.090	50.000	-7.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071442.D **Time Analyzed:** 17:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	571.790	500.000	14.4
Aroclor-1016-2	3.885	3.788	3.988	581.010	500.000	16.2
Aroclor-1016-3	4.046	3.950	4.150	561.130	500.000	12.2
Aroclor-1016-4	4.094	3.998	4.198	548.620	500.000	9.7
Aroclor-1016-5	4.291	4.195	4.395	563.620	500.000	12.7
Aroclor-1260-1	5.292	5.197	5.397	556.780	500.000	11.4
Aroclor-1260-2	5.481	5.386	5.586	560.060	500.000	12.0
Aroclor-1260-3	5.624	5.530	5.730	550.180	500.000	10.0
Aroclor-1260-4	6.089	5.993	6.193	584.560	500.000	16.9
Aroclor-1260-5	6.333	6.238	6.438	562.690	500.000	12.5
Decachlorobiphenyl	7.654	7.559	7.759	52.820	50.000	5.6
Tetrachloro-m-xylene	2.854	2.756	2.956	52.790	50.000	5.6