

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071454.D **Time Analyzed:** 21:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	473.580	500.000	-5.3
Aroclor-1016-2	4.515	4.416	4.616	472.940	500.000	-5.4
Aroclor-1016-3	4.573	4.474	4.674	467.460	500.000	-6.5
Aroclor-1016-4	4.663	4.564	4.764	462.950	500.000	-7.4
Aroclor-1016-5	4.948	4.849	5.049	471.620	500.000	-5.7
Aroclor-1260-1	6.042	5.943	6.143	470.950	500.000	-5.8
Aroclor-1260-2	6.300	6.200	6.400	505.050	500.000	1.0
Aroclor-1260-3	6.652	6.552	6.752	471.430	500.000	-5.7
Aroclor-1260-4	6.867	6.767	6.967	466.390	500.000	-6.7
Aroclor-1260-5	7.175	7.075	7.275	469.930	500.000	-6.0
Decachlorobiphenyl	8.537	8.437	8.637	42.750	50.000	-14.5
Tetrachloro-m-xylene	3.410	3.311	3.511	45.100	50.000	-9.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071454.D **Time Analyzed:** 21:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	477.920	500.000	-4.4
Aroclor-1016-2	3.885	3.788	3.988	487.450	500.000	-2.5
Aroclor-1016-3	4.046	3.950	4.150	522.020	500.000	4.4
Aroclor-1016-4	4.095	3.998	4.198	511.830	500.000	2.4
Aroclor-1016-5	4.293	4.195	4.395	546.060	500.000	9.2
Aroclor-1260-1	5.292	5.197	5.397	454.650	500.000	-9.1
Aroclor-1260-2	5.482	5.386	5.586	458.060	500.000	-8.4
Aroclor-1260-3	5.625	5.530	5.730	441.770	500.000	-11.6
Aroclor-1260-4	6.089	5.993	6.193	483.400	500.000	-3.3
Aroclor-1260-5	6.334	6.238	6.438	461.520	500.000	-7.7
Decachlorobiphenyl	7.653	7.559	7.759	41.190	50.000	-17.6
Tetrachloro-m-xylene	2.855	2.756	2.956	44.840	50.000	-10.3