

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071486.D **Time Analyzed:** 15:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	515.020	500.000	3.0
Aroclor-1016-2	4.516	4.416	4.616	525.670	500.000	5.1
Aroclor-1016-3	4.573	4.474	4.674	519.170	500.000	3.8
Aroclor-1016-4	4.664	4.564	4.764	522.260	500.000	4.5
Aroclor-1016-5	4.948	4.849	5.049	512.840	500.000	2.6
Aroclor-1260-1	6.042	5.943	6.143	477.130	500.000	-4.6
Aroclor-1260-2	6.299	6.200	6.400	501.120	500.000	0.2
Aroclor-1260-3	6.652	6.552	6.752	495.520	500.000	-0.9
Aroclor-1260-4	6.867	6.767	6.967	492.870	500.000	-1.4
Aroclor-1260-5	7.174	7.075	7.275	497.420	500.000	-0.5
Decachlorobiphenyl	8.537	8.437	8.637	42.900	50.000	-14.2
Tetrachloro-m-xylene	3.410	3.311	3.511	46.930	50.000	-6.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071486.D **Time Analyzed:** 15:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.867	3.771	3.971	496.080	500.000	-0.8
Aroclor-1016-2	3.884	3.788	3.988	512.700	500.000	2.5
Aroclor-1016-3	4.045	3.950	4.150	494.830	500.000	-1.0
Aroclor-1016-4	4.094	3.998	4.198	485.710	500.000	-2.9
Aroclor-1016-5	4.291	4.195	4.395	494.130	500.000	-1.2
Aroclor-1260-1	5.291	5.197	5.397	488.180	500.000	-2.4
Aroclor-1260-2	5.481	5.386	5.586	497.580	500.000	-0.5
Aroclor-1260-3	5.624	5.530	5.730	484.620	500.000	-3.1
Aroclor-1260-4	6.088	5.993	6.193	484.460	500.000	-3.1
Aroclor-1260-5	6.332	6.238	6.438	484.670	500.000	-3.1
Decachlorobiphenyl	7.653	7.559	7.759	38.510	50.000	-23.0
Tetrachloro-m-xylene	2.854	2.756	2.956	46.100	50.000	-7.8