

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL07 **Date Analyzed:** 11/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071496.D **Time Analyzed:** 18:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	514.930	500.000	3.0
Aroclor-1016-2	4.516	4.416	4.616	519.110	500.000	3.8
Aroclor-1016-3	4.573	4.474	4.674	518.940	500.000	3.8
Aroclor-1016-4	4.663	4.564	4.764	519.230	500.000	3.8
Aroclor-1016-5	4.949	4.849	5.049	514.700	500.000	2.9
Aroclor-1260-1	6.042	5.943	6.143	485.140	500.000	-3.0
Aroclor-1260-2	6.300	6.200	6.400	510.180	500.000	2.0
Aroclor-1260-3	6.652	6.552	6.752	506.510	500.000	1.3
Aroclor-1260-4	6.866	6.767	6.967	507.470	500.000	1.5
Aroclor-1260-5	7.174	7.075	7.275	516.450	500.000	3.3
Decachlorobiphenyl	8.535	8.437	8.637	44.950	50.000	-10.1
Tetrachloro-m-xylene	3.410	3.311	3.511	46.980	50.000	-6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3674
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL07 **Date Analyzed:** 11/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071496.D **Time Analyzed:** 18:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.867	3.771	3.971	492.920	500.000	-1.4
Aroclor-1016-2	3.884	3.788	3.988	517.450	500.000	3.5
Aroclor-1016-3	4.046	3.950	4.150	500.400	500.000	0.1
Aroclor-1016-4	4.094	3.998	4.198	488.760	500.000	-2.2
Aroclor-1016-5	4.291	4.195	4.395	496.490	500.000	-0.7
Aroclor-1260-1	5.291	5.197	5.397	492.560	500.000	-1.5
Aroclor-1260-2	5.481	5.386	5.586	505.720	500.000	1.1
Aroclor-1260-3	5.624	5.530	5.730	494.120	500.000	-1.2
Aroclor-1260-4	6.088	5.993	6.193	502.120	500.000	0.4
Aroclor-1260-5	6.332	6.238	6.438	504.110	500.000	0.8
Decachlorobiphenyl	7.651	7.559	7.759	41.620	50.000	-16.8
Tetrachloro-m-xylene	2.854	2.756	2.956	46.330	50.000	-7.3