

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB138036
Analysis Date: 11/25/2025
BalanceID: WC SC-5
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 11/25/2025
Extraction IN Time: 11:00
Extraction OUT Time: 11:35
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB138036BL	LB138036BL	WATER	1.3	1000	100	2.7133	2.7133	3.02	2.7134	2.7134	0.0001	0.1
2	LB138036BS	LB138036BS	WATER	1.3	1000	100	2.9102	2.9102	3.01	2.9296	2.9296	0.0194	19.4
3	LB138036BSD	LB138036BSD	WATER	1.3	1000	100	2.7153	2.7153	3.03	2.7352	2.7352	0.0199	19.9
4	Q3675-01	GRAB	WATER	1.6	1000	100	3.1241	3.1241	3.04	3.1524	3.1524	0.0283	28.3
5	Q3705-01	MH-11212025	WATER	1.6	1000	100	3.0256	3.0256	3.05	3.0763	3.0763	0.0507	50.7

QC Batch# LB138036

Test: TPH

Analysis Date: 11/25/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3240
pH Paper 0-14	M6069
Sodium Sulfate	EP2663
1:1 HCL	WP115016
Silica Gel	W3246
Sand	N/A

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP115017
LCSWD	5.00 ML	WP115018
MS/MSD	N/A	N/A

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 13:12

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 12:25

Bal Check Time: 11:11 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 13:45

Out Time1: 13:10

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 15:11

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 14:25

Bal Check Time: 15:39 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 15:37

Out Time2: 15:10

138036

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph q3705 WorkList ID : 193338 Department : Wet-Chemistry Date : 11-25-2025 09:43:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3675-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	E11	11/19/2025	1664A
Q3705-01	MH-1121/2025	Water	TPH	Conc H2SO4 to pH < 2	EURO03	D41	11/21/2025	1664A

Date/Time 11-25-25 10:00
Raw Sample Received by: SP GSC
Raw Sample Relinquished by: CP

Date/Time 11-25-25 15:30
Raw Sample Received by: CP
Raw Sample Relinquished by: SP