

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3675

Project: Monthly 2025

Run No.: LB137969

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137969BS							
BOD5	mg/L	198	170		86	1	84.6-115.4	11/19/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3675

Project: Monthly 2025

Run No.: LB138018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138018BS							
TSS	mg/L	550	591		107	1	90-110	11/21/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3675

Project: Monthly 2025

Run No.: LB138036

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138036BS							
TPH	mg/L	20.0	19.4		97	1	78-114	11/25/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3675

Project: Monthly 2025

Run No.: LB138036

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138036BSD							
TPH	mg/L	20.0	19.9		100	1	78-114	11/25/2025