

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	990	1000	99	95 - 105	P	11/21/2025	13:00	LB138017
	Zinc	1980	2000	99	95 - 105	P	11/21/2025	13:00	LB138017

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.4	20.0	112	80 - 120	P	11/21/2025	13:09	LB138017
	Zinc	44.2	40.0	110	80 - 120	P	11/21/2025	13:09	LB138017

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1260	1250	101	90 - 110	P	11/21/2025	13:44	LB138017
	Zinc	2520	2500	101	90 - 110	P	11/21/2025	13:44	LB138017
CCV02	Copper	1220	1250	98	90 - 110	P	11/21/2025	14:48	LB138017
	Zinc	2440	2500	98	90 - 110	P	11/21/2025	14:48	LB138017
CCV03	Copper	1220	1250	98	90 - 110	P	11/21/2025	15:38	LB138017
	Zinc	2470	2500	99	90 - 110	P	11/21/2025	15:38	LB138017
CCV04	Copper	1220	1250	98	90 - 110	P	11/21/2025	16:28	LB138017
	Zinc	2440	2500	98	90 - 110	P	11/21/2025	16:28	LB138017
CCV05	Copper	1230	1250	98	90 - 110	P	11/21/2025	17:50	LB138017
	Zinc	2420	2500	97	90 - 110	P	11/21/2025	17:50	LB138017
CCV06	Copper	1220	1250	98	90 - 110	P	11/21/2025	18:41	LB138017
	Zinc	2420	2500	97	90 - 110	P	11/21/2025	18:41	LB138017
CCV07	Copper	1220	1250	98	90 - 110	P	11/21/2025	19:31	LB138017
	Zinc	2430	2500	97	90 - 110	P	11/21/2025	19:31	LB138017