



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.8	20.0	109	65 - 135	P	11/21/2025	13:17	LB138017
	Zinc	42.3	40.0	106	65 - 135	P	11/21/2025	13:17	LB138017