



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	11/21/2025	13:13	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	13:13	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	11/21/2025	13:48	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	13:48	LB138017
CCB02	Copper	4.60	+/-10	U	20.0	P	11/21/2025	14:52	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	14:52	LB138017
CCB03	Copper	4.60	+/-10	U	20.0	P	11/21/2025	15:42	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	15:42	LB138017
CCB04	Copper	4.60	+/-10	U	20.0	P	11/21/2025	16:32	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	16:32	LB138017
CCB05	Copper	4.60	+/-10	U	20.0	P	11/21/2025	17:55	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	17:55	LB138017
CCB06	Copper	4.60	+/-10	U	20.0	P	11/21/2025	18:45	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	18:45	LB138017
CCB07	Copper	4.60	+/-10	U	20.0	P	11/21/2025	19:35	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	19:35	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------
