

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	1.60	2.0	80	-18	22	11/21/2025	13:22	LB138017
	Zinc	6.33			-40	40	11/21/2025	13:22	LB138017
ICSAB01	Copper	470	511	92	434	588	11/21/2025	13:26	LB138017
	Zinc	1020	952	107	809	1095	11/21/2025	13:26	LB138017
ICSA	Copper	10.4	2.0	520	-18	22	11/21/2025	13:30	LB138017
	Zinc	14.5			-40	40	11/21/2025	13:30	LB138017
ICSAB	Copper	498	511	98	434	588	11/21/2025	13:34	LB138017
	Zinc	1000	952	105	809	1095	11/21/2025	13:34	LB138017