

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tris Pharma, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q3676</u>
contract: <u>TRIS02</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3676-01</u>	client id: <u>OUTFALL-DSN-001MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3676-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	223		84.5		150	92		P
Zinc	ug/L	75 - 125	115		27.4		100	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tris Pharma, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q3676</u>
contract: <u>TRIS02</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3676-01</u>	client id: <u>OUTFALL-DSN-001MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3676-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	235		84.5		150	101		P
Zinc	ug/L	75 - 125	121		27.4		100	93		P