

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tris Pharma, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3676</u>
Contract:	<u>TRIS02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3676-01</u>	Client ID:	<u>OUTFALL-DSN-001DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3676-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	84.5		82.1		3		P
Zinc	ug/L	20	27.4		17.7	J	43	*	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tris Pharma, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3676</u>
Contract:	<u>TRIS02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3676-01MS</u>	Client ID:	<u>OUTFALL-DSN-001MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3676-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	223		235		5		P
Zinc	ug/L	20	115		121		5		P