

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q3676

Contract: TRIS02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170660BS							
Copper	ug/L	150	159		106	85 - 115	P
Zinc	ug/L	100	109		109	85 - 115	P