

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>TRIS02</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3676</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>11/20/2025 09:13</u>      |
| Lab File ID:        | <u>VN088317.D</u> | Init. Calib. Date(s):  | <u>11/13/2025 11/13/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>10:14 11:50</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND              | RRF   | RRF020 | MIN<br>RRF | %D     | MAX%D |
|-----------------------|-------|--------|------------|--------|-------|
| Ethyl Acetate         | 0.651 | 0.578  |            | -11.21 |       |
| Isopropyl Acetate     | 1.004 | 0.887  |            | -11.65 |       |
| Acetone               | 0.354 | 0.288  |            | -18.64 |       |
| Methylene Chloride    | 2.050 | 1.855  |            | -9.51  |       |
| 1,2-Dichloroethane-d4 | 2.740 | 2.587  | 0.01       | -5.58  |       |
| Toluene-d8            | 1.420 | 1.426  | 0.01       | 0.42   |       |
| 4-Bromofluorobenzene  | 0.507 | 0.477  | 0.2        | -5.92  |       |
| n-amyl Acetate        | 0.509 | 0.447  |            | -12.18 |       |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.