

Report of Analysis

Client: Tris Pharma, Inc.
Project: Quarterly
Client Sample ID: OUTFALL-DSN-001
Lab Sample ID: Q3676-01
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/18/25
Date Received: 11/19/25
SDG No.: Q3676
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
141-78-6	Ethyl Acetate	1.30	U	1	1.30	5.00	ug/L	11/20/25 11:42	VN112025
108-21-4	Isopropyl Acetate	1.00	U	1	1.00	5.00	ug/L	11/20/25 11:42	VN112025
67-64-1	Acetone	280		1	4.60	25.0	ug/L	11/20/25 11:42	VN112025
75-09-2	Methylene Chloride	0.86	U	1	0.86	5.00	ug/L	11/20/25 11:42	VN112025
628-63-7	n-amyl Acetate	0.81	U	1	0.81	5.00	ug/L	11/20/25 11:42	VN112025
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	28.9			91 - 110	96%	SPK: 30		
2037-26-5	Toluene-d8	29.7			91 - 112	99%	SPK: 30		
460-00-4	4-Bromofluorobenzene	24.3			63 - 112	81%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	41400							
540-36-3	1,4-Difluorobenzene	201000							
3114-55-4	Chlorobenzene-d5	178000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products