

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN111325

Review By	John Carlone	Review On	11/14/2025 8:03:53 AM
Supervise By	Maresh Dadoda	Supervise On	11/14/2025 12:39:20 PM
SubDirectory	VN111325	HP Acquire Method	HP Processing Method 624N111325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136189 VP136197,VP136198,VP136199,VP136200,VP136201 VP136212

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088273.D	13 Nov 2025 08:08		JC\MD	Ok
2	VSTDIC005	VSTDIC005	VN088274.D	13 Nov 2025 10:14		JC\MD	Ok,M
3	VSTDICCC020	VSTDICCC020	VN088275.D	13 Nov 2025 10:47		JC\MD	Ok,M
4	VSTDIC050	VSTDIC050	VN088276.D	13 Nov 2025 11:08		JC\MD	Ok,M
5	VSTDIC100	VSTDIC100	VN088277.D	13 Nov 2025 11:29		JC\MD	Ok,M
6	VSTDIC150	VSTDIC150	VN088278.D	13 Nov 2025 11:50		JC\MD	Ok,M
7	IBLK	IBLK	VN088279.D	13 Nov 2025 12:11		JC\MD	Ok
8	VSTDICV020	ICVVN111325	VN088280.D	13 Nov 2025 14:52		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN112025

Review By	John Carlone	Review On	11/21/2025 8:06:56 AM
Supervise By	Amit Patel	Supervise On	11/21/2025 10:08:14 AM
SubDirectory	VN112025	HP Acquire Method	HP Processing Method 624N111325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136340
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136341,VP136342

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088316.D	20 Nov 2025 08:40		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN088317.D	20 Nov 2025 09:13	pH#Lot#V12668	JC\MD	Ok,M
3	VN1120WBS01	VN1120WBS01	VN088318.D	20 Nov 2025 09:49		JC\MD	Ok,M
4	VN1120WBSD01	VN1120WBSD01	VN088319.D	20 Nov 2025 10:22		JC\MD	Ok,M
5	VN1120WBL01	VN1120WBL01	VN088320.D	20 Nov 2025 10:43		JC\MD	Ok
6	Q3689-02	251119025-03 Trip blan	VN088321.D	20 Nov 2025 11:21	vial B pH#6.0 TB	JC\MD	Ok
7	Q3676-01	OUTFALL-DSN-001	VN088322.D	20 Nov 2025 11:42	vial A pH<2	JC\MD	Ok
8	Q3676-04	OUTFALL-DSN-002	VN088323.D	20 Nov 2025 12:02	vial A pH<2 Internal Standard Fail;Need 20X	JC\MD	Dilution
9	Q3676-02MS	OUTFALL-DSN-001MS	VN088324.D	20 Nov 2025 12:23	vial A pH<2	JC\MD	Ok,M
10	Q3676-03MSD	OUTFALL-DSN-001MS	VN088325.D	20 Nov 2025 12:44	vial A pH<2	JC\MD	Ok,M
11	IBLK	IBLK	VN088326.D	20 Nov 2025 13:05		JC\MD	Not Ok
12	Q3689-01	251119071-01 VOA	VN088327.D	20 Nov 2025 13:25	vial B pH#6.0 Surrogate fail - See non con	JC\MD	Ok,M
13	IBLK	IBLK	VN088328.D	20 Nov 2025 13:46		JC\MD	Ok
14	IBLK	IBLK	VN088329.D	20 Nov 2025 14:07		JC\MD	Ok
15	Q3676-04DL	OUTFALL-DSN-002DL	VN088330.D	20 Nov 2025 14:28	vial B pH<2	JC\MD	Ok
16	VSTDCCC020	VSTDCCC020EC	VN088331.D	20 Nov 2025 15:11		JC\MD	Ok,M

M : Manual Integration