

Surrogate Summary

SDG No.: Q3676

Client: Tris Pharma, Inc.

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3676-01	OUTFALL-DSN-001	1,2-Dichloroethane-d4	30	28.9	96		91	110
		Toluene-d8	30	29.7	99		91	112
		4-Bromofluorobenzene	30	24.3	81		63	112
Q3676-02MS	OUTFALL-DSN-001MS	1,2-Dichloroethane-d4	30	30.0	100		91	110
		Toluene-d8	30	30.2	101		91	112
		4-Bromofluorobenzene	30	29.4	98		63	112
Q3676-03MSD	OUTFALL-DSN-001MSD	1,2-Dichloroethane-d4	30	29.0	97		91	110
		Toluene-d8	30	29.3	98		91	112
		4-Bromofluorobenzene	30	28.2	94		63	112
Q3676-04	OUTFALL-DSN-002	1,2-Dichloroethane-d4	30	32.0	107		91	110
		Toluene-d8	30	29.4	98		91	112
		4-Bromofluorobenzene	30	24.6	82		63	112
Q3676-04DL	OUTFALL-DSN-002DL	1,2-Dichloroethane-d4	30	28.7	96		91	110
		Toluene-d8	30	29.8	99		91	112
		4-Bromofluorobenzene	30	26.4	88		63	112
VN1120WBL01	VN1120WBL01	1,2-Dichloroethane-d4	30	28.2	94		91	110
		Toluene-d8	30	29.9	100		91	112
		4-Bromofluorobenzene	30	26.1	87		63	112
VN1120WBS01	VN1120WBS01	1,2-Dichloroethane-d4	30	28.9	96		91	110
		Toluene-d8	30	30.4	101		91	112
		4-Bromofluorobenzene	30	29.1	97		63	112