

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TRIS02
Lab Code: ACE SDG NO.: Q3676
Lab File ID: VN088317.D Date Analyzed: 11/20/2025
Instrument ID: MSVOA_N Time Analyzed: 09:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	49901	7.79	246084	9.08	228625	11.85
UPPER LIMIT	99802	8.294	492168	9.582	457250	12.347
LOWER LIMIT	24950.5	7.294	123042	8.582	114313	11.347
EPA SAMPLE NO.						
OUTFALL-DSN-001	41384	7.80	201252	9.08	178061	11.85
OUTFALL-DSN-001MS	38756	7.80	192300	9.08	177297	11.84
OUTFALL-DSN-001MSD	34515	7.79	167024	9.08	156623	11.85
OUTFALL-DSN-002	18987 *	7.80	97658 *	9.08	86332 *	11.85
OUTFALL-DSN-002DL	49691	7.79	255521	9.08	221473	11.85
VN1120WBL01	53817	7.79	272055	9.08	237385	11.84
VN1120WBS01	59336	7.79	313817	9.08	285258	11.84

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.