

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: Sharkey Landf MA No. : SDG No.: Q3688Initial Calibration Verification Source : Continuing Calibration Verification Source : Run Batch: LB138160 Analytical Method: ICP-AESConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	4000	3900	98	0	5000	5200	104	0	5300	105	0
Barium	8000	8800	110	1	10000	10000	104	0	11000	106	1
Cadmium	2000	2000	102	1	2500	2600	103	1	2600	102	0
Chromium	800	850	106	1	15500	16000	103	0	16000	102	0
Lead	4000	3800	94	0	25000	26000	102	0	26000	102	0
Selenium	4000	4100	102	1	5000	5300	106	1	5400	107	0
Silver	1000	1100	107	1	1250	1300	104	0	1300	102	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: Sharkey Landf MA No. : SDG No.: Q3688Initial Calibration Verification Source : Continuing Calibration Verification Source : Run Batch: LB138160 Analytical Method: ICP-AESConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	5200	105	1			
Barium					10000	11000	107	0			
Cadmium					2500	2500	101	1			
Chromium					15500	15000	100	0			
Lead					25000	25000	100	0			
Selenium					5000	5400	107	1			
Silver					1250	1300	102	0			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: Sharkey Landf MA No. : SDG No.: Q3688
 Initial Calibration Verification Source : MP88222
 Continuing Calibration Verification Source : MP88224
 Run Batch: LB138122 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV069				ID: CCV070		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	96		5.0	5.1	103		5.0	100	