

FORM 6A-OR
GC/MS INITIAL CALIBRATION DATA

Lab Name: Alliance Technical Group, LLC
 Lab Code: ACE Case No.: Sharkey Lar
 Analytical Method : SVOA SIM
 Instrument ID: BNA_N
 GC Column: ZB-GR ID: 0.25 (mm)
 Length : 30 (m)
 Heated Purge: (Y/N)

Contract: 68HERH20D0011
 MA No.: SDG No.: Q3688
 Level : LOW
 Calibration Date(s): 11/26/2025 11/26/2025
 Calibration Time(s): 15:05 18:07
 Purge Volume : (mL)

LAB FILE ID:		RRF0.1 =BN038204.D		RRF0.2= BN038205.D		RRF0.4 =BN038206.D		
		RRF0.8 =BN038207.D		RRF1.6= BN038208.D		RRF3.2 =BN038209.D		
ANALYTE	RRF0.1	RRF0.2	RRF0.4	RRF0.8	RRF1.6	RRF3.2	<u>RRF</u>	% RSD
1,4-Dioxane		0.468	0.460	0.449	0.430	0.453	0.452	3.1
1,4-Dioxane-d8		0.430	0.441	0.428	0.402	0.398	0.420	4.4
Fluoranthene-d10	1.572	1.544	1.619	1.539	1.500		1.555	2.8
2-Methylnaphthalene-d10	0.578	0.574	0.580	0.570	0.551		0.570	2.0