


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112625\
Data File : BN038203.D
Acq On : 26 Nov 2025 14:26
Operator : RC/JU
Sample : DFTPP277
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP277

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN112625.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Dec 01 16:36:13 2025



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.2	236783	PASS
68	69	0.00	2	1.8	4253	PASS
69	198	0.00	100	37.5	238955	PASS
70	69	0.00	2	0.6	1406	PASS
127	198	10	80	47.8	304363	PASS
197	198	0.00	2	0.2	1132	PASS
198	198	100	100	100.0	637013	PASS
199	198	5	9	6.9	43685	PASS
275	198	10	60	27.8	177280	PASS
365	198	1	100	3.4	21539	PASS
441	198	0.01	100	15.0	95331	PASS
442	198	50	100	89.6	571029	PASS
443	442	15	24	18.7	106773	PASS

m/z	Abundance
39.00	274.0
39.90	3184.0
41.00	249.0
42.00	160.0
44.00	4542.0
44.90	199.0
46.90	249.0
48.90	1294.0
49.90	182.0
50.90	465.0
73.00	807.0
77.00	197.0
77.90	185.0
78.80	158.0
83.90	842.0
86.00	541.0
90.80	186.0
95.90	197.0
132.90	305.0
191.00	351.0
192.90	287.0
207.00	2434.0
208.00	702.0
208.90	295.0
281.00	573.0
282.00	281.0
282.90	177.0
355.00	167.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

m/z	Abundance
36.00	321.0
37.00	1346.0
38.00	3005.0
39.00	19520.0
40.00	3688.0
41.00	622.0
42.90	190.0
44.00	5131.0
44.90	500.0
47.00	333.0
47.80	357.0
49.00	2817.0
50.00	66632.0
51.00	256192.0
52.00	12915.0
53.00	750.0
54.90	1266.0
56.00	7708.0
57.00	17216.0
57.90	661.0
60.00	230.0
61.00	3041.0
62.00	4078.0
63.00	9733.0
64.00	1314.0
65.00	5223.0
66.00	254.0
67.00	377.0
68.00	4651.0
69.00	255680.0
70.00	1374.0
70.90	316.0
71.90	256.0
73.00	2673.0
74.00	27488.0
75.00	41048.0
76.00	13586.0
77.00	291072.0
78.00	20944.0
79.00	19216.0
80.00	15515.0
81.00	20288.0
82.00	4751.0
83.00	4962.0
84.00	1207.0
85.00	3323.0
85.90	6318.0
86.90	2913.0
88.00	1092.0
89.00	537.0
91.00	5046.0
92.00	5396.0
93.00	34272.0
94.00	2041.0
95.00	535.0
96.10	1727.0
97.10	468.0
98.00	26640.0
99.00	23848.0
100.00	1841.0
101.00	10957.0
101.90	670.0
102.90	3671.0
104.00	6418.0
105.00	6305.0
106.00	2279.0
107.00	88280.0
108.00	13290.0
109.00	2395.0
110.00	158784.0
111.00	25480.0
112.00	2733.0
113.00	972.0
113.90	324.0
115.00	530.0
116.00	5118.0
117.00	77560.0
118.00	6709.0
118.90	675.0
119.90	1008.0
121.00	468.0
121.90	5642.0
123.00	8846.0
123.90	4567.0
125.00	3673.0
127.00	308672.0
128.00	23776.0
129.00	116400.0
130.00	9584.0
131.00	1762.0
132.00	891.0
133.00	833.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

133.90	3741.0
135.00	8726.0
136.00	3948.0
137.00	6233.0
138.00	925.0
139.00	818.0
140.00	1498.0
141.00	15135.0
142.00	4850.0
143.00	3297.0
144.00	902.0
145.10	935.0
145.90	2971.0
146.90	8293.0
148.00	19576.0
149.00	4545.0
149.90	1174.0
151.00	1970.0
151.70	1151.0
153.00	4661.0
153.90	3810.0
155.00	8562.0
156.00	11920.0
157.10	2189.0
157.90	2638.0
159.00	1810.0
159.90	5455.0
161.00	7089.0
162.00	2085.0
163.00	653.0
163.90	769.0
165.00	6219.0
166.00	4365.0
167.00	31904.0
168.00	23712.0
169.00	3242.0
170.00	762.0
170.90	1117.0
171.90	2709.0
172.90	2962.0
174.00	7079.0
175.00	12919.0
176.00	3286.0
177.00	5527.0
178.00	1518.0
178.90	23560.0
180.00	16560.0
181.00	7017.0
182.00	1082.0
182.90	724.0
183.90	1818.0
185.00	10703.0
186.00	83768.0
187.00	24136.0
188.00	2465.0
188.90	5457.0
189.80	768.0
191.00	2620.0
192.00	7231.0
193.00	7560.0
194.00	1758.0
195.00	916.0
196.00	17640.0
197.00	3396.0
198.00	582976.0
199.00	43152.0
200.00	2942.0
201.50	2763.0
202.90	4223.0
204.00	20704.0
205.00	36096.0
206.00	148480.0
207.00	23248.0
208.00	5026.0
209.00	2095.0
210.00	2347.0
211.00	5966.0
212.90	381.0
213.80	183.0
215.00	1549.0
215.90	3335.0
216.90	36968.0
217.90	4622.0
218.90	538.0
221.00	23880.0
223.00	8405.0
224.00	80288.0
225.00	21688.0
226.00	2752.0
227.00	34616.0
228.00	4929.0
229.00	6360.0
230.00	981.0
231.00	2760.0
231.90	328.0
233.00	557.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

234.00	1861.0
235.00	2293.0
235.90	1471.0
237.00	3805.0
237.90	743.0
239.00	1364.0
240.00	1086.0
240.90	1975.0
242.00	4520.0
243.00	5227.0
244.00	66928.0
245.00	9230.0
246.00	13403.0
246.90	2454.0
248.00	517.0
249.00	2242.0
249.90	675.0
250.90	708.0
251.90	965.0
253.00	1355.0
255.00	329728.0
256.00	47976.0
257.00	4304.0
258.00	20240.0
258.90	3485.0
259.90	571.0
260.90	590.0
264.00	622.0
264.90	7978.0
265.90	1043.0
266.90	254.0
269.90	242.0
270.90	995.0
271.90	704.0
273.00	10592.0
274.00	25680.0
275.00	155136.0
276.00	21480.0
277.00	12365.0
278.00	2601.0
278.90	603.0
281.10	737.0
282.00	661.0
283.00	1390.0
283.90	1132.0
285.00	2185.0
285.90	449.0
289.00	379.0
289.90	714.0
291.00	396.0
292.00	726.0
293.00	2893.0
293.90	932.0
295.00	922.0
295.90	39776.0
297.00	6295.0
298.00	426.0
300.90	486.0
301.90	689.0
303.00	4717.0
304.00	1250.0
304.90	198.0
308.00	555.0
308.90	488.0
310.00	548.0
313.10	400.0
314.00	2353.0
314.90	5227.0
316.00	2432.0
317.00	369.0
319.90	210.0
320.90	1224.0
322.00	705.0
323.00	13357.0
324.00	2435.0
325.10	307.0
325.80	380.0
326.90	2376.0
327.90	1296.0
328.90	213.0
332.00	999.0
333.00	1376.0
334.00	9099.0
335.00	1971.0
336.00	258.0
339.00	264.0
339.90	213.0
341.00	1921.0
341.90	607.0
345.90	3058.0
347.00	548.0
350.90	283.0
352.00	4325.0
353.00	2563.0
354.00	4361.0
355.00	794.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

356.90	211.0
358.90	567.0
364.90	16784.0
365.90	2695.0
366.90	275.0
369.90	351.0
371.00	972.0
372.00	7963.0
373.00	1677.0
373.90	267.0
376.80	162.0
383.00	1940.0
384.00	668.0
384.80	256.0
390.00	804.0
390.90	620.0
392.00	617.0
400.90	490.0
401.90	3385.0
403.00	3852.0
404.00	1396.0
405.00	382.0
421.00	3550.0
421.90	3376.0
423.00	27880.0
424.00	7079.0
424.90	1012.0
441.00	67768.0
442.00	427904.0
443.00	66264.0
444.00	7279.0
445.00	345.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

m/z	Abundance
36.00	379.0
37.00	1235.0
38.10	3392.0
39.00	20432.0
40.00	4045.0
41.00	696.0
43.20	292.0
44.00	5541.0
45.00	718.0
45.90	175.0
46.80	319.0
49.00	3051.0
50.00	74040.0
51.00	280192.0
52.00	13793.0
53.00	589.0
55.00	1570.0
56.00	8212.0
57.00	18568.0
58.00	871.0
59.00	263.0
60.10	302.0
61.00	3501.0
62.00	4300.0
63.00	11139.0
64.00	1316.0
65.00	5425.0
65.90	433.0
67.00	447.0
68.00	5120.0
69.00	280768.0
70.00	1546.0
70.90	249.0
73.00	3032.0
74.00	28536.0
75.00	45600.0
76.00	16464.0
77.00	327872.0
78.00	23016.0
79.00	22320.0
80.00	16592.0
81.00	23928.0
82.00	5441.0
83.00	5352.0
83.90	1420.0
85.00	3400.0
86.00	7148.0
86.90	2928.0
87.90	1574.0
89.00	677.0
91.00	5315.0
91.90	6471.0
93.00	40792.0
94.00	2883.0
95.00	412.0
96.00	1863.0
96.90	698.0
98.00	30680.0
99.00	24368.0
100.00	2281.0
101.00	12874.0
102.00	870.0
102.90	4471.0
104.00	8394.0
105.00	7681.0
106.10	2515.0
107.00	101824.0
108.00	14213.0
110.00	186368.0
111.00	27472.0
111.90	3823.0
112.90	897.0
113.70	266.0
114.20	188.0
114.80	665.0
116.00	5264.0
117.00	92632.0
118.00	7166.0
119.10	943.0
119.90	1494.0
120.90	443.0
122.00	6945.0
123.00	10083.0
124.00	4861.0
125.00	4023.0
127.00	364160.0
128.00	28232.0
129.00	139264.0
130.00	12570.0
131.00	2366.0
132.00	1245.0
132.90	526.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

134.00	4329.0
135.00	11669.0
136.00	4687.0
137.00	6895.0
137.90	1238.0
138.90	719.0
140.00	1598.0
140.90	19192.0
142.00	5633.0
143.00	4542.0
144.00	1243.0
145.00	739.0
145.90	3145.0
147.00	9974.0
148.00	22072.0
148.90	5145.0
150.00	1322.0
151.00	2592.0
151.90	1489.0
153.00	5680.0
154.00	4307.0
155.00	10822.0
156.00	15196.0
157.10	2486.0
158.00	3321.0
159.00	2828.0
159.90	6118.0
161.00	8491.0
162.00	2552.0
163.00	602.0
164.00	1046.0
164.90	7246.0
166.00	5167.0
167.00	40264.0
168.00	25840.0
168.90	3829.0
169.90	1429.0
171.00	1153.0
171.90	3210.0
173.00	4422.0
174.00	8610.0
175.00	14771.0
176.00	4015.0
177.00	6522.0
178.00	2519.0
178.90	28240.0
180.00	19304.0
181.00	10111.0
182.00	1614.0
182.90	798.0
184.00	2307.0
185.00	12360.0
186.00	101792.0
187.00	28936.0
188.00	3425.0
188.90	6468.0
190.00	1042.0
191.00	3537.0
192.00	9690.0
193.00	10256.0
194.00	2045.0
195.10	1144.0
196.00	22288.0
197.90	788864.0
199.00	51184.0
200.00	3544.0
201.50	3279.0
203.00	4887.0
204.00	27096.0
205.00	47680.0
206.00	193792.0
207.00	28880.0
208.00	6259.0
209.00	2551.0
210.10	2869.0
211.00	7542.0
213.00	605.0
213.90	263.0
215.00	2187.0
216.00	4008.0
216.90	53528.0
218.00	6353.0
219.00	534.0
221.00	32408.0
221.70	7395.0
223.00	11709.0
224.00	114992.0
225.00	28440.0
226.00	3092.0
227.00	43568.0
228.00	6222.0
229.00	9983.0
230.00	1432.0
231.00	4074.0
232.00	620.0
233.00	829.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

234.00	2823.0
235.00	3236.0
235.90	2334.0
236.90	4243.0
238.00	580.0
238.90	1535.0
240.00	1824.0
241.00	2575.0
242.00	5453.0
243.00	6900.0
244.00	80376.0
245.00	11966.0
246.00	15520.0
247.00	3039.0
248.00	819.0
248.90	3558.0
249.90	525.0
251.00	682.0
251.90	603.0
253.00	1744.0
255.00	461056.0
256.00	67928.0
257.00	5438.0
257.90	27192.0
259.00	3931.0
259.90	702.0
261.00	716.0
263.00	301.0
264.00	593.0
265.00	10655.0
265.90	1580.0
267.00	401.0
268.00	227.0
270.00	441.0
271.00	1496.0
271.90	1301.0
272.90	13672.0
274.00	36832.0
275.00	214784.0
276.00	29744.0
276.90	18352.0
277.90	3082.0
278.90	636.0
281.00	1151.0
281.90	491.0
283.00	1681.0
284.00	1358.0
285.00	3384.0
286.00	495.0
288.00	350.0
288.90	819.0
289.90	461.0
291.00	557.0
292.10	939.0
293.00	4514.0
294.00	1315.0
295.00	1217.0
296.00	56760.0
297.00	7307.0
297.90	504.0
298.90	154.0
300.90	594.0
302.00	1127.0
303.00	6161.0
304.00	1719.0
305.00	234.0
306.80	160.0
307.90	1157.0
308.90	738.0
310.00	553.0
310.90	183.0
311.90	192.0
313.00	680.0
314.00	2838.0
314.90	7362.0
315.90	3733.0
317.00	717.0
319.90	154.0
321.00	1616.0
322.00	1153.0
323.00	18992.0
324.00	3574.0
324.90	317.0
326.00	492.0
326.90	3958.0
328.00	1975.0
328.90	383.0
329.90	180.0
332.00	1169.0
333.00	2041.0
334.00	14021.0
335.00	3142.0
336.00	451.0
338.90	374.0
340.00	268.0
341.00	2410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

342.00	663.0
343.00	202.0
346.00	5536.0
346.90	1291.0
351.00	551.0
352.00	6484.0
353.00	3940.0
354.00	5779.0
355.00	1748.0
355.90	159.0
359.00	467.0
360.90	178.0
364.90	23920.0
365.90	3803.0
366.80	353.0
369.90	543.0
370.90	1364.0
372.00	9816.0
373.00	2719.0
373.90	311.0
376.90	485.0
383.00	3465.0
383.90	690.0
384.90	415.0
390.00	1548.0
390.90	981.0
391.90	788.0
393.00	165.0
400.90	595.0
401.90	4344.0
403.00	6705.0
404.00	2773.0
405.00	606.0
410.00	236.0
414.90	448.0
421.00	5086.0
422.00	5920.0
423.00	42064.0
424.00	10745.0
425.00	1399.0
437.90	203.0
439.20	287.0
441.00	117552.0
442.00	663424.0
443.00	128576.0
444.00	10536.0
444.90	897.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

m/z	Abundance
36.00	292.0
37.00	821.0
38.00	2038.0
39.00	12392.0
40.00	3545.0
41.00	533.0
42.90	245.0
44.00	5873.0
44.90	444.0
46.90	331.0
48.00	261.0
49.00	2612.0
50.00	46736.0
51.00	175360.0
52.00	8583.0
53.00	531.0
55.00	981.0
56.00	5269.0
57.00	11621.0
58.00	604.0
59.00	295.0
60.00	225.0
61.00	2500.0
62.00	2346.0
63.00	7310.0
64.00	906.0
65.00	3357.0
65.80	237.0
66.90	427.0
68.00	2988.0
69.00	180416.0
69.90	1299.0
73.00	2537.0
74.00	19408.0
75.00	29672.0
76.00	10688.0
77.00	206912.0
78.00	14484.0
79.00	13591.0
80.00	10281.0
81.00	15058.0
82.00	3871.0
83.00	3429.0
83.90	1118.0
85.00	2588.0
86.00	4908.0
87.00	2175.0
88.00	757.0
89.00	437.0
91.00	3376.0
92.00	4004.0
93.00	25616.0
93.90	1732.0
95.00	344.0
96.00	1444.0
97.10	677.0
98.00	19080.0
99.00	16173.0
100.00	1421.0
101.00	8455.0
102.00	514.0
102.90	2732.0
104.00	5619.0
105.00	5272.0
106.00	1827.0
107.00	66568.0
108.00	10467.0
109.00	1910.0
110.00	124840.0
111.00	18176.0
112.00	2013.0
112.90	781.0
114.80	320.0
115.90	3881.0
117.00	63296.0
118.00	5234.0
119.00	819.0
120.00	839.0
121.00	433.0
122.00	4743.0
123.00	6923.0
124.00	3113.0
125.00	2990.0
127.00	240256.0
128.00	18704.0
129.00	92760.0
130.00	7733.0
130.90	1382.0
132.00	916.0
133.00	657.0
134.00	2550.0
134.90	7568.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

136.00	3124.0
137.00	4374.0
138.00	870.0
138.90	450.0
139.90	1169.0
141.00	12025.0
142.00	3847.0
143.00	2863.0
143.90	634.0
144.90	711.0
145.90	2373.0
146.90	6658.0
147.90	15584.0
149.00	3658.0
150.00	841.0
151.00	1481.0
151.90	968.0
153.00	4014.0
153.90	2824.0
155.00	6975.0
156.00	10188.0
157.00	1967.0
158.00	2268.0
159.00	1530.0
160.00	4046.0
161.00	6093.0
161.90	1492.0
162.90	455.0
164.00	664.0
165.00	4961.0
165.90	3933.0
167.00	27696.0
168.00	17704.0
169.00	2329.0
169.90	791.0
170.90	1097.0
171.90	2162.0
173.00	3000.0
174.00	4861.0
175.00	10016.0
176.00	2893.0
177.00	4702.0
177.90	1700.0
178.90	19656.0
180.00	13675.0
181.00	5991.0
182.00	1081.0
182.90	776.0
183.90	1597.0
185.00	9341.0
186.00	72832.0
187.00	20528.0
188.00	2264.0
188.90	4890.0
189.90	746.0
190.90	2141.0
191.90	6370.0
193.00	7409.0
194.00	1408.0
195.10	742.0
196.00	14755.0
197.90	539200.0
198.90	36720.0
200.00	2474.0
201.60	2190.0
202.90	3771.0
204.00	19552.0
205.00	29368.0
206.00	128896.0
207.00	19368.0
208.00	4503.0
209.00	1758.0
210.00	2316.0
211.00	5520.0
212.90	412.0
214.00	173.0
214.90	1472.0
216.00	3177.0
216.90	35760.0
218.00	4768.0
218.90	285.0
221.00	23112.0
221.80	4855.0
223.00	8411.0
224.00	78304.0
225.00	19528.0
226.00	2320.0
227.00	32472.0
228.00	5187.0
229.00	7050.0
229.90	853.0
231.00	2878.0
232.00	548.0
232.90	792.0
233.90	1972.0
235.00	2152.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

236.00	1449.0
237.00	3245.0
238.00	610.0
239.00	1476.0
239.90	965.0
240.90	1775.0
241.90	4447.0
243.00	4731.0
244.00	67584.0
245.00	8910.0
246.00	12468.0
247.00	2403.0
248.00	525.0
248.90	2466.0
249.90	531.0
250.90	626.0
252.00	599.0
252.90	1386.0
255.00	320000.0
256.00	50344.0
257.00	3828.0
258.00	17792.0
258.90	2921.0
259.90	685.0
260.90	425.0
263.90	407.0
265.00	8185.0
265.90	1088.0
267.00	227.0
270.00	341.0
271.00	985.0
272.00	1002.0
273.00	9685.0
274.00	29976.0
275.00	161920.0
276.00	22456.0
277.00	12136.0
278.00	2434.0
278.90	523.0
281.00	1110.0
281.90	564.0
283.00	1229.0
284.00	947.0
285.00	2464.0
285.90	334.0
288.00	288.0
288.90	499.0
289.90	366.0
290.90	462.0
292.00	651.0
292.90	3048.0
293.90	941.0
296.00	47120.0
297.00	7394.0
298.00	376.0
298.90	205.0
301.00	690.0
302.00	870.0
303.00	5194.0
303.90	1327.0
305.00	204.0
308.00	699.0
309.00	401.0
309.90	830.0
310.90	173.0
312.80	266.0
313.90	2252.0
314.90	4747.0
316.00	2869.0
317.00	431.0
319.90	194.0
320.90	1636.0
322.00	670.0
323.00	17168.0
324.00	3077.0
325.10	314.0
325.90	394.0
327.00	2815.0
327.90	1599.0
328.90	222.0
332.00	1158.0
332.90	1579.0
334.00	9587.0
335.00	2203.0
335.90	264.0
338.80	269.0
340.00	303.0
341.00	2307.0
341.90	477.0
346.00	3834.0
346.90	793.0
351.00	471.0
352.00	5292.0
353.00	3518.0
353.90	5032.0
354.90	1050.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277

357.80	192.0
358.90	297.0
364.90	23912.0
365.90	3048.0
366.80	335.0
370.00	802.0
371.00	1440.0
371.90	7979.0
373.00	2475.0
373.80	210.0
376.90	175.0
383.00	2256.0
383.90	715.0
385.00	212.0
389.90	1064.0
391.10	791.0
392.00	609.0
401.00	601.0
402.00	3466.0
403.00	5756.0
404.00	2619.0
405.00	353.0
409.90	271.0
414.90	367.0
421.00	4815.0
422.00	4975.0
422.90	31480.0
424.10	8655.0
424.90	1008.0
437.90	251.0
439.30	314.0
441.00	100672.0
442.00	621760.0
443.00	125480.0
444.10	12488.0
445.10	558.0

Instrument :
BNA_N
ClientSampleId :
DFTPP277