

FORM 2A-OR
DEUTERATED MONITORING COMPOUND RECOVERY

Lab Name : Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code : ACE Case No.: Sharkey L& MA No.: _____ SDG No.: Q3688
 Analytical Method : SVOA SIM Level : _____
 Matrix : Water

EPA Sample No .	DMC1 (FLN)	DMC2 (2MN)	DMC3 (DXE)	DMC4	DMC5	DMC6	DMC7	DMC8	DMC9
FDGW-20251120	90	67	56						
M-28I	97	88	82						
M-28S	87	85	66						
M-28SDL	83	88	32						
M-29	101	93	77						
M-30I	93	65	68						
M-30S	80	55	62						
M-31I	110	66	66						
M-31S	94	54	54						
M-31SMS	90	49	47						
M-31SMSD	84	49	50						
M-32I	117	65	63						
M-32S	102	71	66						
RBGW-20251119	87	58	63						
RBGW-20251120	95	70	64						
RBGW-20251121	97	85	68						
RBGW-20251124	100	93	60						
SBLK718	72	79	79						
SBLK720	95	89	81						
SBLK722	100	94	85						
SBLK724	80	86	91						
SBLK742	77	84	84						
SLCS718	81	88	97						
SLCS720	68	81	96						
SLCS722	72	86	104						
SLCS724	59	69	78						
SLCS742	76	95	104						
WS-1	87	80	63						
WS-14	106	95	70						

QC LIMITS

DMC1 (FLN) = Fluoranthene-d10 (SIM)	(30-130)
DMC2 (2MN) = 2-Methylnapthalene-d10 (SIM)	(30-130)
DMC3 (DXE) = 1,4-Dioxane-d8 (SIM)	(40-110)