

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG112425

Review By	Jagrut Upadhyay	Review On	11/25/2025 6:00:52 PM
Supervise By	mohammad ahmed	Supervise On	12/3/2025 1:12:53 AM
SubDirectory	BG112425	HP Acquire Method	BNA_G
		HP Processing Method	BG112425
STD. NAME	STD REF.#		
Tune/Reschk	SP6868		
Initial Calibration Stds	SP6887,SP6888,SP6889,SP6890,SP6891,SP6892		
CCC	SP6889		
Internal Standard/PEM	S13184,10ul/1000ul sample		
ICV/I.BLK	SP6930		
Surrogate Standard	SP6932		
MS/MSD Standard	SP6931		
LCS Standard	SP6931		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP460	BG064847.D	24 Nov 2025 13:23	RC/JU	Ok
2	SSTD00519	BG064848.D	24 Nov 2025 14:04	RC/JU	Ok,M
3	SSTD01020	BG064849.D	24 Nov 2025 14:45	RC/JU	Ok,M
4	SSTD02021	BG064850.D	24 Nov 2025 15:26	RC/JU	Ok,M
5	SSTD04022	BG064851.D	24 Nov 2025 16:06	RC/JU	Ok,M
6	SSTD08023	BG064852.D	24 Nov 2025 16:46	RC/JU	Ok,M
7	SSTD16024	BG064853.D	24 Nov 2025 17:27	RC/JU	Ok,M
8	SSTDICV020	BG064854.D	24 Nov 2025 18:08	RC/JU	Ok,M

M : Manual Integration

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG112525

Review By	Jagrut Upadhyay	Review On	12/2/2025 3:50:25 PM
Supervise By	mohammad ahmed	Supervise On	12/3/2025 1:12:48 AM
SubDirectory	BG112525	HP Acquire Method	BNA_G
		HP Processing Method	BG112425
STD. NAME	STD REF.#		
Tune/Reschk	SP6868		
Initial Calibration Stds	SP6887,SP6888,SP6889,SP6890,SP6891,SP6892		
CCC	SP6889		
Internal Standard/PEM	S13184,10ul/1000ul sample		
ICV/I.BLK	SP6930		
Surrogate Standard	SP6932		
MS/MSD Standard	SP6931		
LCS Standard	SP6931		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP461	BG064855.D	25 Nov 2025 10:29	RC/JU	Ok
2	SSTDCCC020	BG064856.D	25 Nov 2025 11:10	RC/JU	Ok,M
3	PB170717BL	BG064857.D	25 Nov 2025 12:28	RC/JU	Ok
4	PB170717BS	BG064858.D	25 Nov 2025 13:09	RC/JU	Ok,M
5	Q3688-01	BG064859.D	25 Nov 2025 13:52	RC/JU	Ok,M
6	Q3688-02MS	BG064860.D	25 Nov 2025 14:33	RC/JU	Ok,M
7	Q3688-03MSD	BG064861.D	25 Nov 2025 15:14	RC/JU	Ok,M
8	Q3688-04	BG064862.D	25 Nov 2025 15:55	RC/JU	Ok,M
9	Q3688-06	BG064863.D	25 Nov 2025 16:36	RC/JU	Ok,M
10	Q3688-09	BG064864.D	25 Nov 2025 17:17	RC/JU	Ok,M
11	Q3688-10	BG064865.D	25 Nov 2025 17:58	RC/JU	Ok,M
12	Q3688-11	BG064866.D	25 Nov 2025 18:38	RC/JU	Ok,M
13	PB170719BS	BG064867.D	25 Nov 2025 19:19	RC/JU	Ok,M
14	PB170719BL	BG064868.D	25 Nov 2025 20:00	RC/JU	Ok,M
15	SSTDCCC020	BG064869.D	25 Nov 2025 21:22	RC/JU	Ok,M
16	PB170721BL	BG064870.D	25 Nov 2025 22:03	RC/JU	Not Ok
17	PB170723BL	BG064871.D	25 Nov 2025 22:44	RC/JU	Not Ok
18	Q3688-12	BG064872.D	25 Nov 2025 23:25	RC/JU	Not Ok
19	Q3688-13	BG064873.D	26 Nov 2025 00:05	RC/JU	Not Ok
20	Q3688-15	BG064874.D	26 Nov 2025 00:46	RC/JU	Not Ok
21	Q3688-16	BG064875.D	26 Nov 2025 1:27	RC/JU	Not Ok

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG112525

Review By	Jagrut Upadhyay	Review On	12/2/2025 3:50:25 PM
Supervise By	mohammad ahmed	Supervise On	12/3/2025 1:12:48 AM
SubDirectory	BG112525	HP Acquire Method	BNA_G
		HP Processing Method	BG112425
STD. NAME	STD REF.#		
Tune/Reschk	SP6868		
Initial Calibration Stds	SP6887,SP6888,SP6889,SP6890,SP6891,SP6892		
CCC	SP6889		
Internal Standard/PEM	S13184,10ul/1000ul sample		
ICV/I.BLK	SP6930		
Surrogate Standard	SP6932		
MS/MSD Standard	SP6931		
LCS Standard	SP6931		

22	Q3688-17	BG064876.D	26 Nov 2025 2:08	RC/JU	Not Ok
23	Q3688-18	BG064877.D	26 Nov 2025 2:49	RC/JU	Not Ok
24	Q3688-19	BG064878.D	26 Nov 2025 3:30	RC/JU	Not Ok
25	Q3688-20	BG064879.D	26 Nov 2025 4:11	RC/JU	Not Ok
26	PB170721BS	BG064880.D	26 Nov 2025 4:52	RC/JU	Not Ok
27	PB170723BS	BG064881.D	26 Nov 2025 5:32	RC/JU	Not Ok
28	SSTDCCC020	BG064882.D	26 Nov 2025 6:54	RC/JU	Ok,M
29	PB170721BL	BG064883.D	26 Nov 2025 10:54	RC/JU	Ok
30	PB170723BL	BG064884.D	26 Nov 2025 11:35	RC/JU	Ok
31	Q3688-12	BG064885.D	26 Nov 2025 12:16	RC/JU	Ok,M
32	Q3688-13	BG064886.D	26 Nov 2025 12:57	RC/JU	Ok,M
33	Q3688-15	BG064887.D	26 Nov 2025 13:38	RC/JU	Ok,M
34	Q3688-16	BG064888.D	26 Nov 2025 14:19	RC/JU	Ok,M
35	Q3688-17	BG064889.D	26 Nov 2025 15:00	RC/JU	Ok,M
36	SSTDCCC020	BG064890.D	26 Nov 2025 15:42	RC/JU	Ok,M
37	PB170741BL	BG064891.D	26 Nov 2025 16:23	RC/JU	Ok,M
38	PB170721BS	BG064892.D	26 Nov 2025 17:04	RC/JU	Ok,M
39	PB170723BS	BG064893.D	26 Nov 2025 17:45	RC/JU	Ok,M
40	PB170741BS	BG064894.D	26 Nov 2025 18:25	RC/JU	Ok,M
41	Q3688-18	BG064895.D	26 Nov 2025 19:06	RC/JU	Ok,M
42	Q3688-19	BG064896.D	26 Nov 2025 19:47	RC/JU	Ok,M
43	Q3688-20	BG064897.D	26 Nov 2025 20:28	RC/JU	Ok,M
44	Q3688-22	BG064898.D	26 Nov 2025 21:09	RC/JU	Ok,M

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG112525

Review By	Jagrut Upadhyay	Review On	12/2/2025 3:50:25 PM
Supervise By	mohammad ahmed	Supervise On	12/3/2025 1:12:48 AM
SubDirectory	BG112525	HP Acquire Method	BNA_G
		HP Processing Method	BG112425
STD. NAME	STD REF.#		
Tune/Reschk	SP6868		
Initial Calibration Stds	SP6887,SP6888,SP6889,SP6890,SP6891,SP6892		
CCC	SP6889		
Internal Standard/PEM	S13184,10ul/1000ul sample		
ICV/I.BLK	SP6930		
Surrogate Standard	SP6932		
MS/MSD Standard	SP6931		
LCS Standard	SP6931		

45	Q3688-23	BG064899.D	26 Nov 2025 21:50	RC/JU	Ok,M
46	SSTDCCC020EC	BG064900.D	26 Nov 2025 22:31	RC/JU	Ok,M

M : Manual Integration