

FORM 2B-OR
DEUTERATED MONITORING COMPOUND RECOVERY

Lab Name : Alliance Technical Group, LLC Contract : 68HERH20D0011
 Lab Code : ACE Case No.: Sharkey Lar MA No.: _____ SDG No.: Q3688
 Analytical Method : VOA Level : _____
 Matrix : Water

EPA Sample No .	DMC9 (TOL)	DMC10 (TDP)	DMC11 (HEX)	DMC12 (TCA)	DMC13 (DCZ)	DMC14	DMC15	DMC16	TOT OUT
VBLK130	93	85	93	96	101				0
RBGW-20251119	95	86	86	96	100				0
TBGW-20251119	88	80	83	91	96				0
M-32I	89	79	83	90	97				0
RBGW-20251120	93	85	87	94	100				0
TBGW-20251120	97	88	89	96	103				0
FDGW-20251120	89	81	85	91	98				0
M-31S	87	76	81	88	95				0
M-31I	85	77	82	87	94				0
M-32S	92	79	87	95	99				0
M-30S	93	83	87	92	100				0
M-30I	90	80	85	91	98				0
M-31SMS	90	86	95	95	97				0
M-31SMSD	90	84	94	95	92				0
VBLK132	93	90	88	97	99				0
RBGW-20251121	95	93	92	102	106				0
TBGW-20251121	90	86	85	95	99				0
RBGW-20251124	93	87	89	99	103				0
TBGW-20251124	86	82	84	91	97				0
WS-14	95	87	86	100	103				0
M-29	92	84	84	95	101				0
M-28I	94	87	87	101	106				0
WS-1	93	86	84	99	106				0
M-285	89	84	83	95	99				0
VHBLK001	97	86	96	104	109				0

QC LIMITS

DMC9 (TOL) = Toluene-d8	(80-120)
DMC10 (TDP) = trans-1,3-Dichloropropene-d4	(60-125)
DMC11 (HEX) = 2-Hexanone-d5	(45-130)
DMC12 (TCA) = 1,1,2,2-Tetrachloroethane-d2	(65-120)
DMC13 (DCZ) = 1,2-Dichlorobenzene-d4	(80-120)