

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : WASTE WATER 2025

GARDEN STATE LABORATORIES, INC.

410 Hillside Avenue

Hillside, NJ - 07205

Phone No: 800-273-8901

ORDER ID : Q3689

ATTENTION : Sharon Ercoliani



Laboratory Certification ID # 20012



Table Of Contents for Q3689

1) Signature Page	3
2) Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) VOCMS Group1 Data	8
6) VOCMS Group2 Data	13
7) Shipping Document	23
7.1) CHAIN OF CUSTODY	24
7.2) Lab Certificate	25
7.3) Internal COC	26

Cover Page

Order ID : Q3689

Project ID : Waste Water 2025

Client : Garden State Laboratories, Inc.

Lab Sample Number

Q3689-01
Q3689-02

Client Sample Number

251119071-01 VOA
251119025-03 Trip blank

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Sohil Jodhani, QA/QC Director at 1:28 pm, Nov 26, 2025

Date: 11/25/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Garden State Laboratories, Inc.

Project Name: Waste Water 2025

Project # N/A

Order ID # Q3689

Test Name: VOCMS Group1,VOCMS Group2

A. Number of Samples and Date of Receipt:

2 Water samples were received on 11/20/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group1,VOCMS Group2. This data package contains results for VOCMS Group1(624.1),VOCMS Group2(8260-Low).

C. Analytical Techniques:

VOCMS Group1 : The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOCMS Group1 was based on method 624.1.

VOCMS Group2 : The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group2 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except following VOCMS Group1 : 251119071-01 VOA [1,2-Dichloroethane-d4 - 111%], %, due to difficult matrix, therefore no corrective action taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD recoveries met criteria.

The Blank Spike met requirements for all compounds except following

VOCMS Group2 : The Blank Spike for {VX1120WBS01} with File ID: VX048626.D met requirements for all compounds except for 1,1,1-Trichloroethane[112%], 1,1,2-Trichloroethane[128%], 1,1,2-Trichlorotrifluoroethane[123%], 1,1-Dichloroethene[112%], 1,2-Dibromoethane[123%], 1,2-Dichloropropane[115%], 2-Hexanone[120%], 4-Methyl-2-Pentanone[120%], Bromodichloromethane[120%], Bromomethane[127%], cis-1,3-Dichloropropene[119%], Dibromochloromethane[125%], Isopropylbenzene[117%], t-1,3-Dichloropropene[115%] and Toluene[123%].

284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

The Blank Spike Duplicate met requirements for all compounds except following VOCMS Group2 : The Blank Spike Duplicate for {VX1120WBSD01} with File ID: VX048630.D met requirements for all compounds except for 1,1,2-Trichloroethane[141%], 1,2-Dibromoethane[134%], 2-Hexanone[130%], 4-Methyl-2-Pentanone[130%], Bromodichloromethane[120%], cis-1,3-Dichloropropene[117%], Dibromochloromethane[136%], t-1,3-Dichloropropene[113%] and Toluene[114%]. Above Blank Spike and Blank Spike Duplicate failing high for above mentioned compound and associate sample # 01 having positive hit for some compounds, but now no more vial left for reanalysis therefore lab reported this analysis as final.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements except following VOCMS Group2 : The Continuous Calibration File ID VX048623.D met the requirements except for Carbon Disulfide failing marginally low therefore no corrective action taken.

The Tuning criteria met requirements.

VOCMS Group2 : Sample 251119071-01 VOA was diluted due to high concentration.

E. Additional Comments:

VOCMS Group1 : “As per method, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project. However, Lab has performed LCS/LCSD instead.”

VOCMS Group2 : Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____

APPROVED

By Sohil Jodhani, QA/QC Director at 1:28 pm, Nov 26, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3689

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MOHAMMAD AHMED

Date: 11/25/2025

Hit Summary Sheet

624.1

SDG No.: Q3689

Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:

A

B

C

D



SAMPLE DATA

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119071-01 VOA
Lab Sample ID: Q3689-01
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
107-02-8	Acrolein	6.60	U	1	6.60	25.0	ug/L	11/20/25 13:25	VN112025
107-13-1	Acrylonitrile	2.80	U	1	2.80	25.0	ug/L	11/20/25 13:25	VN112025
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	33.2	*		91 - 110	111%	SPK: 30		
2037-26-5	Toluene-d8	31.2			91 - 112	104%	SPK: 30		
460-00-4	4-Bromofluorobenzene	27.5			63 - 112	92%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	27600							
540-36-3	1,4-Difluorobenzene	217000							
3114-55-4	Chlorobenzene-d5	176000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119025-03 Trip blank
Lab Sample ID: Q3689-02
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
107-02-8	Acrolein	6.60	U	1	6.60	25.0	ug/L	11/20/25 11:21	VN112025
107-13-1	Acrylonitrile	2.80	U	1	2.80	25.0	ug/L	11/20/25 11:21	VN112025
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	28.0			91 - 110	93%	SPK: 30		
2037-26-5	Toluene-d8	29.3			91 - 112	98%	SPK: 30		
460-00-4	4-Bromofluorobenzene	23.6			63 - 112	79%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	46600							
540-36-3	1,4-Difluorobenzene	225000							
3114-55-4	Chlorobenzene-d5	206000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3689	OrderDate:	11/20/2025 9:02:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3689-01	251119071-01 VOA	Water	VOCMS Group1	624.1	11/19/25		11/20/25	11/20/25
Q3689-02	251119025-03 Trip blank	Water	VOCMS Group1	624.1	11/19/25		11/20/25	11/20/25

Hit Summary Sheet
8260-Low

SDG No.: Q3689
Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: 251119071-01 VOA								
Q3689-01	251119071-01 VOA Water		Vinyl Chloride	0.48	J	0.26	1.00	ug/L
Q3689-01	251119071-01 VOA Water		Acetone	960	E	1.50	5.00	ug/L
Q3689-01	251119071-01 VOA Water		Methyl Acetate	2.60		0.27	1.00	ug/L
Q3689-01	251119071-01 VOA Water		Benzene	3.70		0.15	1.00	ug/L
Q3689-01	251119071-01 VOA Water		4-Methyl-2-Pentanone	14.1	Q	0.68	5.00	ug/L
Q3689-01	251119071-01 VOA Water		Toluene	21.2	Q	0.14	1.00	ug/L
Q3689-01	251119071-01 VOA Water		2-Hexanone	3.40	JQ	0.89	5.00	ug/L
Q3689-01	251119071-01 VOA Water		Chlorobenzene	1.60		0.12	1.00	ug/L
Q3689-01	251119071-01 VOA Water		Ethyl Benzene	8.30		0.13	1.00	ug/L
Q3689-01	251119071-01 VOA Water		m/p-Xylenes	8.50		0.24	2.00	ug/L
Q3689-01	251119071-01 VOA Water		o-Xylene	4.70		0.12	1.00	ug/L
Q3689-01	251119071-01 VOA Water		Isopropylbenzene	1.10	Q	0.12	1.00	ug/L
Q3689-01	251119071-01 VOA Water		1,4-Dichlorobenzene	2.70		0.19	1.00	ug/L
Q3689-01	251119071-01 VOA Water		1,2-Dichlorobenzene	0.21	J	0.16	1.00	ug/L
			Total Voc :			1030		
Q3689-01	251119071-01 VOA Water		2-Pentanone	* 13.7	J	0	0	ug/L
Q3689-01	251119071-01 VOA Water		(+)-2-Bornanone	* 120	J	0	0	ug/L
Q3689-01	251119071-01 VOA Water		Eucalyptol	* 15.3	J	0	0	ug/L
Q3689-01	251119071-01 VOA Water		3-Pentanone, 2,4-dimethyl-	* 14.2	J	0	0	ug/L
Q3689-01	251119071-01 VOA Water		Fenchone	* 63.2	J	0	0	ug/L
Q3689-01	251119071-01 VOA Water		Tetrahydrofuran	* 790	J	0.99	5.00	ug/L
Q3689-01	251119071-01 VOA Water		Tert butyl alcohol	* 9200	J	5.50	25.0	ug/L
Q3689-01	251119071-01 VOA Water		Diethyl Ether	* 6.00	J	0.31	1.00	ug/L
Q3689-01	251119071-01 VOA Water		n-propylbenzene	* 0.41	J	0.13	1.00	ug/L
Q3689-01	251119071-01 VOA Water		1,3,5-Trimethylbenzene	* 0.76	J	0.15	1.00	ug/L
Q3689-01	251119071-01 VOA Water		1,2,4-Trimethylbenzene	* 3.00	J	0.14	1.00	ug/L
Q3689-01	251119071-01 VOA Water		p-Isopropyltoluene	* 6.70	J	0.13	1.00	ug/L
Q3689-01	251119071-01 VOA Water		Naphthalene	* 23.0	J	0.20	1.00	ug/L
			Total Tics :			10300		
			Total Concentration:			11300		
Client ID: 251119071-01 VOADL								
Q3689-01DL	251119071-01 VOA Water		Acetone	910	D	30.2	100	ug/L
Q3689-01DL	251119071-01 VOA Water		Benzene	4.80	JD	3.00	20.0	ug/L
Q3689-01DL	251119071-01 VOA Water		4-Methyl-2-Pentanone	14.1	JDQ	13.6	100	ug/L
Q3689-01DL	251119071-01 VOA Water		Toluene	23.7	DQ	2.80	20.0	ug/L
Q3689-01DL	251119071-01 VOA Water		Ethyl Benzene	8.70	JD	2.60	20.0	ug/L

Hit Summary Sheet
8260-Low

SDG No.: Q3689

Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3689-01DL	251119071-01	VOA Water	m/p-Xylenes	9.70	JD	4.80	40.0	ug/L
Q3689-01DL	251119071-01	VOA Water	o-Xylene	5.40	JD	2.40	20.0	ug/L
Total Voc :				976				
Total Concentration:				976				

A

B

C

D



SAMPLE DATA

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119071-01 VOA
Lab Sample ID: Q3689-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/20/25 13:48	VX112025
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/20/25 13:48	VX112025
75-01-4	Vinyl Chloride	0.48	J	1	0.26	1.00	ug/L	11/20/25 13:48	VX112025
74-83-9	Bromomethane	1.40	UQ	1	1.40	5.00	ug/L	11/20/25 13:48	VX112025
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/20/25 13:48	VX112025
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/20/25 13:48	VX112025
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	UQ	1	0.25	1.00	ug/L	11/20/25 13:48	VX112025
75-35-4	1,1-Dichloroethene	0.23	UQ	1	0.23	1.00	ug/L	11/20/25 13:48	VX112025
67-64-1	Acetone	960	E	1	1.50	5.00	ug/L	11/20/25 13:48	VX112025
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/20/25 13:48	VX112025
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:48	VX112025
79-20-9	Methyl Acetate	2.60		1	0.27	1.00	ug/L	11/20/25 13:48	VX112025
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/20/25 13:48	VX112025
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/20/25 13:48	VX112025
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/20/25 13:48	VX112025
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/20/25 13:48	VX112025
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/20/25 13:48	VX112025
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/20/25 13:48	VX112025
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/20/25 13:48	VX112025
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/20/25 13:48	VX112025
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/20/25 13:48	VX112025
71-55-6	1,1,1-Trichloroethane	0.20	UQ	1	0.20	1.00	ug/L	11/20/25 13:48	VX112025
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:48	VX112025
71-43-2	Benzene	3.70		1	0.15	1.00	ug/L	11/20/25 13:48	VX112025
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/20/25 13:48	VX112025
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/20/25 13:48	VX112025
78-87-5	1,2-Dichloropropane	0.20	UQ	1	0.20	1.00	ug/L	11/20/25 13:48	VX112025
75-27-4	Bromodichloromethane	0.22	UQ	1	0.22	1.00	ug/L	11/20/25 13:48	VX112025
108-10-1	4-Methyl-2-Pentanone	14.1	Q	1	0.68	5.00	ug/L	11/20/25 13:48	VX112025
108-88-3	Toluene	21.2	Q	1	0.14	1.00	ug/L	11/20/25 13:48	VX112025
10061-02-6	t-1,3-Dichloropropene	0.17	UQ	1	0.17	1.00	ug/L	11/20/25 13:48	VX112025
10061-01-5	cis-1,3-Dichloropropene	0.16	UQ	1	0.16	1.00	ug/L	11/20/25 13:48	VX112025
79-00-5	1,1,2-Trichloroethane	0.21	UQ	1	0.21	1.00	ug/L	11/20/25 13:48	VX112025
591-78-6	2-Hexanone	3.40	JQ	1	0.89	5.00	ug/L	11/20/25 13:48	VX112025
124-48-1	Dibromochloromethane	0.18	UQ	1	0.18	1.00	ug/L	11/20/25 13:48	VX112025
106-93-4	1,2-Dibromoethane	0.15	UQ	1	0.15	1.00	ug/L	11/20/25 13:48	VX112025
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/20/25 13:48	VX112025
108-90-7	Chlorobenzene	1.60		1	0.12	1.00	ug/L	11/20/25 13:48	VX112025
100-41-4	Ethyl Benzene	8.30		1	0.13	1.00	ug/L	11/20/25 13:48	VX112025
179601-23-1	m/p-Xylenes	8.50		1	0.24	2.00	ug/L	11/20/25 13:48	VX112025
95-47-6	o-Xylene	4.70		1	0.12	1.00	ug/L	11/20/25 13:48	VX112025
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/20/25 13:48	VX112025
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/20/25 13:48	VX112025

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119071-01 VOA
Lab Sample ID: Q3689-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	1.10	Q	1	0.12	1.00	ug/L	11/20/25 13:48	VX112025
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/20/25 13:48	VX112025
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:48	VX112025
106-46-7	1,4-Dichlorobenzene	2.70		1	0.19	1.00	ug/L	11/20/25 13:48	VX112025
95-50-1	1,2-Dichlorobenzene	0.21	J	1	0.16	1.00	ug/L	11/20/25 13:48	VX112025
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/20/25 13:48	VX112025
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/20/25 13:48	VX112025
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/20/25 13:48	VX112025
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	58.6			74 - 125	117%	SPK: 50		
1868-53-7	Dibromofluoromethane	42.8			75 - 124	86%	SPK: 50		
2037-26-5	Toluene-d8	49.4			86 - 113	99%	SPK: 50		
460-00-4	4-Bromofluorobenzene	60.2			77 - 121	120%	SPK: 50		
INTERNAL STANDARDS									
363-72-4	Pentafluorobenzene	146000							
540-36-3	1,4-Difluorobenzene	300000							
3114-55-4	Chlorobenzene-d5	324000							
3855-82-1	1,4-Dichlorobenzene-d4	180000							
TENTATIVE IDENTIFIED COMPOUNDS									
60-29-7	Diethyl Ether	6.00	J			2.14	ug/L		
75-65-0	Tert butyl alcohol	9200	J			2.93	ug/L		
109-99-9	Tetrahydrofuran	790	J			4.98	ug/L		
000107-87-9	2-Pentanone	13.7	J			7.44	ug/L		
000565-80-0	3-Pentanone, 2,4-dimethyl-	14.2	J			9.36	ug/L		
103-65-1	n-propylbenzene	0.41	J			11.3	ug/L		
108-67-8	1,3,5-Trimethylbenzene	0.76	J			11.4	ug/L		
95-63-6	1,2,4-Trimethylbenzene	3.00	J			11.7	ug/L		
99-87-6	p-Isopropyltoluene	6.70	J			12.0	ug/L		
000470-82-6	Eucalyptol	15.3	J			12.1	ug/L		
001195-79-5	Fenchone	63.2	J			12.9	ug/L		
000464-49-3	(+)-2-Bornanone	120	J			13.5	ug/L		
91-20-3	Naphthalene	23.0	J			13.8	ug/L		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119071-01 VOADL
Lab Sample ID: Q3689-01DL
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	4.40	UD	20	4.40	20.0	ug/L	11/20/25 14:55	VX112025
74-87-3	Chloromethane	6.40	UD	20	6.40	20.0	ug/L	11/20/25 14:55	VX112025
75-01-4	Vinyl Chloride	5.20	UD	20	5.20	20.0	ug/L	11/20/25 14:55	VX112025
74-83-9	Bromomethane	28.8	UDQ20		28.8	100	ug/L	11/20/25 14:55	VX112025
75-00-3	Chloroethane	9.40	UD	20	9.40	20.0	ug/L	11/20/25 14:55	VX112025
75-69-4	Trichlorofluoromethane	6.60	UD	20	6.60	20.0	ug/L	11/20/25 14:55	VX112025
76-13-1	1,1,2-Trichlorotrifluoroethane	5.00	UDQ20		5.00	20.0	ug/L	11/20/25 14:55	VX112025
75-35-4	1,1-Dichloroethene	4.60	UDQ20		4.60	20.0	ug/L	11/20/25 14:55	VX112025
67-64-1	Acetone	910	D	20	30.2	100	ug/L	11/20/25 14:55	VX112025
75-15-0	Carbon Disulfide	4.20	UD	20	4.20	20.0	ug/L	11/20/25 14:55	VX112025
1634-04-4	Methyl tert-butyl Ether	3.20	UD	20	3.20	20.0	ug/L	11/20/25 14:55	VX112025
79-20-9	Methyl Acetate	5.40	UD	20	5.40	20.0	ug/L	11/20/25 14:55	VX112025
75-09-2	Methylene Chloride	5.60	UD	20	5.60	20.0	ug/L	11/20/25 14:55	VX112025
156-60-5	trans-1,2-Dichloroethene	4.60	UD	20	4.60	20.0	ug/L	11/20/25 14:55	VX112025
75-34-3	1,1-Dichloroethane	4.60	UD	20	4.60	20.0	ug/L	11/20/25 14:55	VX112025
110-82-7	Cyclohexane	29.0	UD	20	29.0	100	ug/L	11/20/25 14:55	VX112025
78-93-3	2-Butanone	19.6	UD	20	19.6	100	ug/L	11/20/25 14:55	VX112025
56-23-5	Carbon Tetrachloride	5.00	UD	20	5.00	20.0	ug/L	11/20/25 14:55	VX112025
156-59-2	cis-1,2-Dichloroethene	3.80	UD	20	3.80	20.0	ug/L	11/20/25 14:55	VX112025
74-97-5	Bromochloromethane	4.40	UD	20	4.40	20.0	ug/L	11/20/25 14:55	VX112025
67-66-3	Chloroform	5.00	UD	20	5.00	20.0	ug/L	11/20/25 14:55	VX112025
71-55-6	1,1,1-Trichloroethane	4.00	UDQ20		4.00	20.0	ug/L	11/20/25 14:55	VX112025
108-87-2	Methylcyclohexane	3.20	UD	20	3.20	20.0	ug/L	11/20/25 14:55	VX112025
71-43-2	Benzene	4.80	JD	20	3.00	20.0	ug/L	11/20/25 14:55	VX112025
107-06-2	1,2-Dichloroethane	4.40	UD	20	4.40	20.0	ug/L	11/20/25 14:55	VX112025
79-01-6	Trichloroethene	1.90	UD	20	1.90	20.0	ug/L	11/20/25 14:55	VX112025
78-87-5	1,2-Dichloropropane	4.00	UDQ20		4.00	20.0	ug/L	11/20/25 14:55	VX112025
75-27-4	Bromodichloromethane	4.40	UDQ20		4.40	20.0	ug/L	11/20/25 14:55	VX112025
108-10-1	4-Methyl-2-Pentanone	14.1	JDQ	20	13.6	100	ug/L	11/20/25 14:55	VX112025
108-88-3	Toluene	23.7	DQ	20	2.80	20.0	ug/L	11/20/25 14:55	VX112025
10061-02-6	t-1,3-Dichloropropene	3.40	UDQ20		3.40	20.0	ug/L	11/20/25 14:55	VX112025
10061-01-5	cis-1,3-Dichloropropene	3.20	UDQ20		3.20	20.0	ug/L	11/20/25 14:55	VX112025
79-00-5	1,1,2-Trichloroethane	4.20	UDQ20		4.20	20.0	ug/L	11/20/25 14:55	VX112025
591-78-6	2-Hexanone	17.8	UDQ20		17.8	100	ug/L	11/20/25 14:55	VX112025
124-48-1	Dibromochloromethane	3.60	UDQ20		3.60	20.0	ug/L	11/20/25 14:55	VX112025
106-93-4	1,2-Dibromoethane	3.00	UDQ20		3.00	20.0	ug/L	11/20/25 14:55	VX112025
127-18-4	Tetrachloroethene	4.60	UD	20	4.60	20.0	ug/L	11/20/25 14:55	VX112025
108-90-7	Chlorobenzene	2.40	UD	20	2.40	20.0	ug/L	11/20/25 14:55	VX112025
100-41-4	Ethyl Benzene	8.70	JD	20	2.60	20.0	ug/L	11/20/25 14:55	VX112025
179601-23-1	m/p-Xylenes	9.70	JD	20	4.80	40.0	ug/L	11/20/25 14:55	VX112025
95-47-6	o-Xylene	5.40	JD	20	2.40	20.0	ug/L	11/20/25 14:55	VX112025
100-42-5	Styrene	3.00	UD	20	3.00	20.0	ug/L	11/20/25 14:55	VX112025
75-25-2	Bromoform	3.80	UD	20	3.80	20.0	ug/L	11/20/25 14:55	VX112025

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119071-01 VOADL
Lab Sample ID: Q3689-01DL
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	2.40	UDQ20	2.40		20.0	ug/L	11/20/25 14:55	VX112025
79-34-5	1,1,2,2-Tetrachloroethane	5.20	UD 20	5.20		20.0	ug/L	11/20/25 14:55	VX112025
541-73-1	1,3-Dichlorobenzene	3.20	UD 20	3.20		20.0	ug/L	11/20/25 14:55	VX112025
106-46-7	1,4-Dichlorobenzene	3.80	UD 20	3.80		20.0	ug/L	11/20/25 14:55	VX112025
95-50-1	1,2-Dichlorobenzene	3.20	UD 20	3.20		20.0	ug/L	11/20/25 14:55	VX112025
96-12-8	1,2-Dibromo-3-Chloropropane	10.6	UD 20	10.6		20.0	ug/L	11/20/25 14:55	VX112025
120-82-1	1,2,4-Trichlorobenzene	4.00	UD 20	4.00		20.0	ug/L	11/20/25 14:55	VX112025
87-61-6	1,2,3-Trichlorobenzene	4.00	UD 20	4.00		20.0	ug/L	11/20/25 14:55	VX112025
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	57.6		74 - 125		115%	SPK: 50		
1868-53-7	Dibromofluoromethane	43.8		75 - 124		88%	SPK: 50		
2037-26-5	Toluene-d8	50.8		86 - 113		102%	SPK: 50		
460-00-4	4-Bromofluorobenzene	60.3		77 - 121		121%	SPK: 50		
INTERNAL STANDARDS									
363-72-4	Pentafluorobenzene	151000							
540-36-3	1,4-Difluorobenzene	295000							
3114-55-4	Chlorobenzene-d5	326000							
3855-82-1	1,4-Dichlorobenzene-d4	176000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119025-03 Trip blank
Lab Sample ID: Q3689-02
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/20/25 13:07	VX112025
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/20/25 13:07	VX112025
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	11/20/25 13:07	VX112025
74-83-9	Bromomethane	1.40	UQ	1	1.40	5.00	ug/L	11/20/25 13:07	VX112025
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/20/25 13:07	VX112025
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/20/25 13:07	VX112025
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	UQ	1	0.25	1.00	ug/L	11/20/25 13:07	VX112025
75-35-4	1,1-Dichloroethene	0.23	UQ	1	0.23	1.00	ug/L	11/20/25 13:07	VX112025
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	11/20/25 13:07	VX112025
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/20/25 13:07	VX112025
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:07	VX112025
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	11/20/25 13:07	VX112025
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/20/25 13:07	VX112025
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/20/25 13:07	VX112025
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/20/25 13:07	VX112025
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/20/25 13:07	VX112025
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/20/25 13:07	VX112025
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/20/25 13:07	VX112025
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/20/25 13:07	VX112025
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/20/25 13:07	VX112025
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/20/25 13:07	VX112025
71-55-6	1,1,1-Trichloroethane	0.20	UQ	1	0.20	1.00	ug/L	11/20/25 13:07	VX112025
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:07	VX112025
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	11/20/25 13:07	VX112025
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/20/25 13:07	VX112025
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/20/25 13:07	VX112025
78-87-5	1,2-Dichloropropane	0.20	UQ	1	0.20	1.00	ug/L	11/20/25 13:07	VX112025
75-27-4	Bromodichloromethane	0.22	UQ	1	0.22	1.00	ug/L	11/20/25 13:07	VX112025
108-10-1	4-Methyl-2-Pentanone	0.68	UQ	1	0.68	5.00	ug/L	11/20/25 13:07	VX112025
108-88-3	Toluene	0.14	UQ	1	0.14	1.00	ug/L	11/20/25 13:07	VX112025
10061-02-6	t-1,3-Dichloropropene	0.17	UQ	1	0.17	1.00	ug/L	11/20/25 13:07	VX112025
10061-01-5	cis-1,3-Dichloropropene	0.16	UQ	1	0.16	1.00	ug/L	11/20/25 13:07	VX112025
79-00-5	1,1,2-Trichloroethane	0.21	UQ	1	0.21	1.00	ug/L	11/20/25 13:07	VX112025
591-78-6	2-Hexanone	0.89	UQ	1	0.89	5.00	ug/L	11/20/25 13:07	VX112025
124-48-1	Dibromochloromethane	0.18	UQ	1	0.18	1.00	ug/L	11/20/25 13:07	VX112025
106-93-4	1,2-Dibromoethane	0.15	UQ	1	0.15	1.00	ug/L	11/20/25 13:07	VX112025
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/20/25 13:07	VX112025
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	11/20/25 13:07	VX112025
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	11/20/25 13:07	VX112025
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	11/20/25 13:07	VX112025
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	11/20/25 13:07	VX112025
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/20/25 13:07	VX112025
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/20/25 13:07	VX112025

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: 251119025-03 Trip blank
Lab Sample ID: Q3689-02
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/19/25
Date Received: 11/20/25
SDG No.: Q3689
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.12	UQ	1	0.12	1.00	ug/L	11/20/25 13:07	VX112025
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/20/25 13:07	VX112025
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:07	VX112025
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	11/20/25 13:07	VX112025
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/20/25 13:07	VX112025
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/20/25 13:07	VX112025
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/20/25 13:07	VX112025
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/20/25 13:07	VX112025
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	56.5			74 - 125	113%	SPK: 50		
1868-53-7	Dibromofluoromethane	47.3			75 - 124	95%	SPK: 50		
2037-26-5	Toluene-d8	47.9			86 - 113	96%	SPK: 50		
460-00-4	4-Bromofluorobenzene	56.5			77 - 121	113%	SPK: 50		
INTERNAL STANDARDS									
363-72-4	Pentafluorobenzene	189000							
540-36-3	1,4-Difluorobenzene	351000							
3114-55-4	Chlorobenzene-d5	372000							
3855-82-1	1,4-Dichlorobenzene-d4	196000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3689	OrderDate:	11/20/2025 9:02:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3689-01	251119071-01 VOA	Water			11/19/25			11/20/25
			VOCMS Group1	624.1			11/20/25	
			VOCMS Group2	8260-Low			11/20/25	
Q3689-01DL	251119071-01 VOADL	Water			11/19/25			11/20/25
			VOCMS Group2	8260-Low			11/20/25	
Q3689-02	251119025-03 Trip blank	Water			11/19/25			11/20/25
			VOCMS Group1	624.1			11/20/25	
			VOCMS Group2	8260-Low			11/20/25	



SHIPPING DOCUMENTS

Garden State Laboratories, Inc.

Main Lab - 410 Hillside Avenue, Hillside NJ 07205 - NJDEP Lab Cert. #20044
Jersey Shore Lab - 54 Main Street, Waretown NJ 08758 - NJDEP Lab Cert. #15037

Tel. 800-273-8901/908-688-8900 Fax 908-688-8966 www.gslabs.com info@gslabs.com

Office and Drop off Locations

North Jersey Office: 225 Sparta Avenue, Sparta, NJ 07871 Tel. 973-729-1827

West Jersey Office: 2050 Route 31 North, Glen Gardner, NJ 08826 Tel. 908-537-7414

CLIENT INFORMATION (REPORT TO BE SENT TO)

Name: Garden State Laboratories, Inc.

Contact/Authorized by: Robert Szot

Mailing Address: 410 Hillside Ave.

Phone: 908-688-8900 EXT 129

City/State/Zip: Hillside, NJ. 07205

Email: rszot@gslabs.com

SAMPLE INFORMATION

SAMPLE TYPE: WASTE WATER

SAMPLE LOCATION: ACUA SW LANDFILL LEACHATE TANKS

FOR SAMPLE RECEIVING USE ONLY

DATE/TIME/TEMP. REC'D AT LAB:

Page _____ of _____

GSL CLIENT

MICRO #

CHEM. #

SAMPLE REC'D BY:

☐ GSL FIELD SAMPLER/PICK-UP

☐ PICK-UP AT DROP OFF LOCATION

☐ DELIVERED BY CLIENT

Grab Comp	SAMPLE ID	SAMPLE COLLECTION				ANALYSIS REQUIRED (Print Legibly)	CONTAINER INFORMATION			
		Date	Time	AM	PM		No.	Type*	Size	Pres.*
☒	351119071-01 VOA	11/19/25	8:56			EPA 8260 TCL LIST + Acrolien & Acrylonitrile	3	V	40mL	A
☒	351119025-03 Trip blank	-	-	-	-	EPA 8260 TCL LIST + Acrolien & Acrylonitrile	2	V	40mL	A

*Container Type: P = Plastic G = Glass A = Amber Glass I = Sterile Thio V = Vial Other/Specify:

*Preservation Code: A = Non Preserved B = Sulfuric Acid C = Sodium Hydroxide D = Nitric Acid
 E = Hydrochloric Acid F = Zinc Acetate G = Sodium Thiosulfate H = Ascorbic Acid I = Cooled Other/Specify:

☒ SUBCONTRACTED WORK

TURNAROUND TIME: ☒ Standard ☐ Rush (IF RUSH REQUESTED) Rush Due by:

SEND TO: Chem Tech

REPORT FORMAT: ☒ Standard Report ☐ Other/Specify:

DATE/TIME:

☐ Standard Report + E2 PWS ID#:

METHOD OF SHIPMENT:

PAYMENT INFORMATION

Deliver

☐ Sampling/Pick-up Fee: \$

☐ Composite Fee: \$

☐ Rush Fee: \$

Amount Due: \$

Payment Method: ☐ Credit Card Type:

☐ Check #

Other: See Quote

Note:

VOA UNPRESERVED DUE TO EFFERVESCENCE - 3 DAY TAT PER JORDAN HEDVAT AT216

SAMPLE CUSTODY EXCHANGES MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION
PLEASE PRINT YOUR NAME LEGIBLY, USE FULL LEGAL SIGNATURE, DATE AND TIME

Sampled by (PRINT):

Signature:

Date/Time:

Client/Client's Representative (PRINT):

Signature:

Date/Time:

1. Received/Relinquished by (PRINT): Kaylee Evans

Signature: Kaylee Evans

Date/Time: 11/19/25 15:06

2. Received/Relinquished by (PRINT): Luis Melendez

Signature: Luis Melendez

Date/Time: 11/19/25 8:57

The liability of Garden State Laboratories, Inc. for services rendered shall in no event exceed the amount of the invoice.
 Main Lab certified by NJ Dept. of Health, NJDEP-TNII, NY Dept. of Health #11550 and PADEP #68-03880

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3689	GARD04	Order Date : 11/20/2025 9:02:00 AM	Project Mgr :
Client Name : Garden State Laboratories, I		Project Name : Waste Water 2025	Report Type : Level 1
Client Contact : Sharon Ercoliani		Receive DateTime : 11/20/2025 8:57:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Garden State Laboratories, I		Purchase Order :	Hard Copy Date :
Invoice Contact : Sharon Ercoliani			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3689-01	251119071-01 VOA	Water	11/19/2025	08:56					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	
Q3689-02	251119025-03 Trip blank	Water	11/19/2025	08:56					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time :

CP
11/20/25 9:30

Received By :

Date / Time :

[Signature]
11/20/25

Storage Area : VOA Refridgerator Room

930