

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GARD04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3689</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>11/20/2025 09:13</u>
Lab File ID:	<u>VN088317.D</u>	Init. Calib. Date(s):	<u>11/13/2025 11/13/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:14 11:50</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.392	0.304		-22.45	
Acrylonitrile	1.209	0.978		-19.11	
1,2-Dichloroethane-d4	2.740	2.587	0.01	-5.58	
Toluene-d8	1.420	1.426	0.01	0.42	
4-Bromofluorobenzene	0.507	0.477	0.2	-5.92	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.