

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04  
 Lab Code: ACE SDG NO.: Q3689  
 Lab File ID: VX048623.D Date Analyzed: 11/20/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 08:51  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                         | IS4<br>AREA # | RT #   |  |  |  |  |
|-------------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD             | 140728        | 12.00€ |  |  |  |  |
| UPPER LIMIT             | 281456        | 12.50€ |  |  |  |  |
| LOWER LIMIT             | 70364         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO.          |               |        |  |  |  |  |
| 251119071-01 VOA        | 179606        | 12.01  |  |  |  |  |
| 251119071-01 VOADL      | 176431        | 12.01  |  |  |  |  |
| 251119025-03 Trip blank | 196397        | 12.01  |  |  |  |  |
| VX1120WBL01             | 223366        | 12.01  |  |  |  |  |
| VX1120WBS01             | 116936        | 12.01  |  |  |  |  |
| VX1120WBSD01            | 109343        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.