

### Duplicate Sample Summary

|                   |                               |                                         |          |
|-------------------|-------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Tetra Tech NUS, Inc.          | <b>SDG No.:</b>                         | Q3690    |
| <b>Project:</b>   | NWIRP Bethpage 112G08005-WE13 | <b>Sample ID:</b>                       | Q3676-04 |
| <b>Client ID:</b> | OUTFALL-DSN-002DUP            | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| TSS     | mg/L  | +/-5                | 2890             |                    | 2890                |                    | 1                  | 0.05       |      | 11/21/2025       |

## Duplicate Sample Summary

|                                               |                                           |
|-----------------------------------------------|-------------------------------------------|
| <b>Client:</b> Tetra Tech NUS, Inc.           | <b>SDG No.:</b> Q3690                     |
| <b>Project:</b> NWIRP Bethpage 112G08005-WE13 | <b>Sample ID:</b> Q3690-01                |
| <b>Client ID:</b> RW8-SP100-20251118DUP       | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| TDS     | mg/L  | +/-5                | 2.00             | J                  | 2.00                | J                  | 1                  | 0          |      | 11/20/2025       |