

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3690

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3840	4000	96	90 - 110	P	11/21/2025	13:00	LB138017

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3690

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	101	100	101	80 - 120	P	11/21/2025	13:09	LB138017

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3690

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5020	5000	100	90 - 110	P	11/21/2025	13:44	LB138017
CCV02	Iron	4800	5000	96	90 - 110	P	11/21/2025	14:48	LB138017
CCV03	Iron	4870	5000	98	90 - 110	P	11/21/2025	15:38	LB138017
CCV04	Iron	4860	5000	97	90 - 110	P	11/21/2025	16:28	LB138017
CCV05	Iron	4860	5000	97	90 - 110	P	11/21/2025	17:50	LB138017
CCV06	Iron	4950	5000	99	90 - 110	P	11/21/2025	18:41	LB138017
CCV07	Iron	4950	5000	99	90 - 110	P	11/21/2025	19:31	LB138017