

metals  
- 5a -  
**MATRIX SPIKE SUMMARY**

**client:** Tetra Tech NUS, Inc.      **level:** low      **sdg no.:** Q3690  
**contract:** TETR06      **lab code:** ACE  
**matrix:** Water      **sample id:** Q3687-01      **client id:** N43502MS  
**Percent Solids for Sample:** NA      **Spiked ID:** Q3687-01MS      **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Iron    | ug/L  | 87 - 115               | 11800            |   | 10700            |   | 1500           | 73            |      | P |

metals  
- 5a -  
**MATRIX SPIKE DUPLICATE SUMMARY**

**client:** Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q3690  
**contract:** TETR06 **lab code:** ACE  
**matrix:** Water **sample id:** Q3687-01 **client id:** N43502MSD  
**Percent Solids for Sample:** NA **Spiked ID:** Q3687-01MSD **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit %R | MSD<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|---------------|---|------------------|---|----------------|---------------|------|---|
| Iron    | ug/L  | 87 - 115               | 12300         |   | 10700            |   | 1500           | 109           |      | P |