

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3690

Client ID : SSTDCCC0.4

Date Analyzed: 12/01/2025

Lab File ID: BN038230.D

Time Analyzed: 17:51

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8054	7.768	24560	10.56	13292	14.43
UPPER LIMIT	16108	8.268	49120	11.062	26584	14.93
LOWER LIMIT	4027	7.268	12280	10.062	6646	13.93
EPA SAMPLE NO.						
01 PB170712BS	7901	7.77	22379	10.56	10358	14.43
02 PB170712BL	8256	7.77	22952	10.57	10779	14.43
03 PB170712BSD	9081	7.77	25072	10.56	10824	14.43
04 RW8-SP100-20251118	9532	7.77	28839	10.56	13442	14.42
05 RW8-SP303-20251118	8417	7.77	23771	10.56	11088	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.