

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3690

Client ID: SSTDCCC0.4

Date Analyzed: 12/01/2025

Lab File ID: BN038230.D

Time Analyzed: 17:51

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	21033	17.173	11703	21.366	11756	23.674
UPPER LIMIT	42066	17.673	23406	21.866	23512	24.174
LOWER LIMIT	10516.5	16.673	5851.5	20.866	5878	23.174
EPA SAMPLE NO.						
01 PB170712BS	14739	17.19	10142	21.37	10375	23.67
02 PB170712BL	14873	17.19	10479	21.38	10853	23.68
03 PB170712BSD	16101	17.17	14003	21.37	15858	23.67
04 RW8-SP100-20251118	20421	17.17	15899	21.37	16933	23.67
05 RW8-SP303-20251118	18387	17.17	14715	21.37	15551	23.67

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.