

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3692
Instrument ID: BNA_N Calibration Date/Time: 12/01/2025 17:51
Lab File ID: BN038230.D Init. Calib. Date(s): 11/26/2025 11/27/2025
EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 20:35 00:13
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.568	0.591		4.0	20.0
Fluoranthene-d10	0.868	0.851		-2.0	20.0
2-Fluorophenol	0.935	0.944		1.0	20.0
Phenol-d6	1.168	1.204		3.1	20.0
Nitrobenzene-d5	0.320	0.317		-0.9	20.0
2-Fluorobiphenyl	1.545	1.656		7.2	20.0
2,4,6-Tribromophenol	0.126	0.139		10.3	20.0
Terphenyl-d14	0.924	0.961		4.0	20.0
1,4-Dioxane	0.421	0.451		7.1	20.0

All other compounds must meet a minimum RRF of 0.010.