

Surrogate Summary

SW-846

SDG No.: Q3692

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170712BL	PB170712BL	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.36	91		30	150
		Nitrobenzene-d5	0.4	0.36	90		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.35	88		58	132
PB170712BS	PB170712BS	2-Methylnaphthalene-d10	0.4	0.41	101		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.40	99		55	111
		2-Fluorobiphenyl	0.4	0.45	113	*	53	106
		Terphenyl-d14	0.4	0.36	91		58	132
PB170712BSD	PB170712BSD	2-Methylnaphthalene-d10	0.4	0.35	86		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.40	100		55	111
		2-Fluorobiphenyl	0.4	0.43	107	*	53	106
		Terphenyl-d14	0.4	0.32	79		58	132
Q3692-01	RW5-SP100-20251118	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.44	109		30	150
		Nitrobenzene-d5	0.4	0.35	86		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.43	108		58	132
Q3692-02	RW5-SP201-20251118	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.43	106		30	150
		Nitrobenzene-d5	0.4	0.38	95		55	111
		2-Fluorobiphenyl	0.4	0.44	109	*	53	106
		Terphenyl-d14	0.4	0.45	113		58	132
Q3692-03	RW5-SP303-20251118	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.45	112		30	150
		Nitrobenzene-d5	0.4	0.35	87		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.40	100		58	132