

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3693
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071466.D **Time Analyzed:** 01:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	495.860	500.000	-0.8
Aroclor-1016-2	4.515	4.416	4.616	496.710	500.000	-0.7
Aroclor-1016-3	4.572	4.474	4.674	494.030	500.000	-1.2
Aroclor-1016-4	4.663	4.564	4.764	494.140	500.000	-1.2
Aroclor-1016-5	4.948	4.849	5.049	493.390	500.000	-1.3
Aroclor-1260-1	6.042	5.943	6.143	502.410	500.000	0.5
Aroclor-1260-2	6.300	6.200	6.400	533.840	500.000	6.8
Aroclor-1260-3	6.653	6.552	6.752	497.560	500.000	-0.5
Aroclor-1260-4	6.867	6.767	6.967	493.730	500.000	-1.3
Aroclor-1260-5	7.175	7.075	7.275	498.260	500.000	-0.3
Decachlorobiphenyl	8.534	8.437	8.637	44.880	50.000	-10.2
Tetrachloro-m-xylene	3.409	3.311	3.511	47.860	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3693
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071466.D **Time Analyzed:** 01:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.867	3.771	3.971	495.010	500.000	-1.0
Aroclor-1016-2	3.883	3.788	3.988	479.710	500.000	-4.1
Aroclor-1016-3	4.045	3.950	4.150	481.800	500.000	-3.6
Aroclor-1016-4	4.093	3.998	4.198	481.250	500.000	-3.8
Aroclor-1016-5	4.291	4.195	4.395	479.140	500.000	-4.2
Aroclor-1260-1	5.291	5.197	5.397	483.670	500.000	-3.3
Aroclor-1260-2	5.481	5.386	5.586	476.900	500.000	-4.6
Aroclor-1260-3	5.624	5.530	5.730	468.150	500.000	-6.4
Aroclor-1260-4	6.088	5.993	6.193	509.780	500.000	2.0
Aroclor-1260-5	6.333	6.238	6.438	486.030	500.000	-2.8
Decachlorobiphenyl	7.652	7.559	7.759	43.290	50.000	-13.4
Tetrachloro-m-xylene	2.853	2.756	2.956	47.240	50.000	-5.5