

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3693
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071499.D **Time Analyzed:** 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.497	4.397	4.597	500.080	500.000	0.0
Aroclor-1016-2	4.516	4.416	4.616	511.920	500.000	2.4
Aroclor-1016-3	4.573	4.474	4.674	516.030	500.000	3.2
Aroclor-1016-4	4.663	4.564	4.764	499.380	500.000	-0.1
Aroclor-1016-5	4.949	4.849	5.049	502.550	500.000	0.5
Aroclor-1260-1	6.043	5.943	6.143	472.510	500.000	-5.5
Aroclor-1260-2	6.299	6.200	6.400	496.740	500.000	-0.7
Aroclor-1260-3	6.653	6.552	6.752	491.380	500.000	-1.7
Aroclor-1260-4	6.868	6.767	6.967	495.770	500.000	-0.8
Aroclor-1260-5	7.175	7.075	7.275	499.050	500.000	-0.2
Decachlorobiphenyl	8.538	8.437	8.637	43.350	50.000	-13.3
Tetrachloro-m-xylene	3.411	3.311	3.511	46.830	50.000	-6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3693
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071499.D **Time Analyzed:** 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.867	3.771	3.971	480.120	500.000	-4.0
Aroclor-1016-2	3.884	3.788	3.988	497.630	500.000	-0.5
Aroclor-1016-3	4.046	3.950	4.150	491.710	500.000	-1.7
Aroclor-1016-4	4.094	3.998	4.198	471.270	500.000	-5.7
Aroclor-1016-5	4.291	4.195	4.395	478.620	500.000	-4.3
Aroclor-1260-1	5.291	5.197	5.397	472.130	500.000	-5.6
Aroclor-1260-2	5.481	5.386	5.586	483.690	500.000	-3.3
Aroclor-1260-3	5.624	5.530	5.730	473.140	500.000	-5.4
Aroclor-1260-4	6.088	5.993	6.193	482.510	500.000	-3.5
Aroclor-1260-5	6.333	6.238	6.438	485.120	500.000	-3.0
Decachlorobiphenyl	7.652	7.559	7.759	40.100	50.000	-19.8
Tetrachloro-m-xylene	2.854	2.756	2.956	45.000	50.000	-10.0