

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3693
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL07 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071511.D **Time Analyzed:** 13:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	551.680	500.000	10.3
Aroclor-1016-2	4.515	4.416	4.616	558.150	500.000	11.6
Aroclor-1016-3	4.573	4.474	4.674	543.710	500.000	8.7
Aroclor-1016-4	4.663	4.564	4.764	547.330	500.000	9.5
Aroclor-1016-5	4.948	4.849	5.049	550.300	500.000	10.1
Aroclor-1260-1	6.042	5.943	6.143	517.660	500.000	3.5
Aroclor-1260-2	6.299	6.200	6.400	544.480	500.000	8.9
Aroclor-1260-3	6.652	6.552	6.752	541.520	500.000	8.3
Aroclor-1260-4	6.866	6.767	6.967	545.400	500.000	9.1
Aroclor-1260-5	7.173	7.075	7.275	549.740	500.000	9.9
Decachlorobiphenyl	8.536	8.437	8.637	47.490	50.000	-5.0
Tetrachloro-m-xylene	3.410	3.311	3.511	51.050	50.000	2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3693
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL07 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071511.D **Time Analyzed:** 13:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.867	3.771	3.971	545.530	500.000	9.1
Aroclor-1016-2	3.883	3.788	3.988	553.600	500.000	10.7
Aroclor-1016-3	4.045	3.950	4.150	538.380	500.000	7.7
Aroclor-1016-4	4.094	3.998	4.198	528.700	500.000	5.7
Aroclor-1016-5	4.291	4.195	4.395	536.720	500.000	7.3
Aroclor-1260-1	5.291	5.197	5.397	528.660	500.000	5.7
Aroclor-1260-2	5.480	5.386	5.586	540.430	500.000	8.1
Aroclor-1260-3	5.624	5.530	5.730	529.360	500.000	5.9
Aroclor-1260-4	6.087	5.993	6.193	533.010	500.000	6.6
Aroclor-1260-5	6.332	6.238	6.438	535.550	500.000	7.1
Decachlorobiphenyl	7.652	7.559	7.759	44.920	50.000	-10.2
Tetrachloro-m-xylene	2.854	2.756	2.956	50.910	50.000	1.8