

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114869.D	04 Nov 2025 09:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO114870.D	04 Nov 2025 09:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114872.D	04 Nov 2025 09:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114873.D	04 Nov 2025 10:16		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114874.D	04 Nov 2025 10:34		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO114875.D	04 Nov 2025 10:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114876.D	04 Nov 2025 11:11		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114877.D	04 Nov 2025 11:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114879.D	04 Nov 2025 12:06		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114880.D	04 Nov 2025 12:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114881.D	04 Nov 2025 12:42		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114882.D	04 Nov 2025 13:00		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114884.D	04 Nov 2025 13:37		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114885.D	04 Nov 2025 13:56		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114886.D	04 Nov 2025 14:14		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114887.D	04 Nov 2025 14:33		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114889.D	04 Nov 2025 15:09		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114890.D	04 Nov 2025 15:28		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114891.D	04 Nov 2025 16:04		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114892.D	04 Nov 2025 16:23		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO114893.D	04 Nov 2025 17:00		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114895.D	04 Nov 2025 17:36		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114896.D	04 Nov 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114897.D	04 Nov 2025 18:13		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114898.D	04 Nov 2025 18:31		YP/AJ	Ok
31	PO110425ICV500	ICVPO110425	PO114899.D	04 Nov 2025 19:08		YP/AJ	Ok
32	AR1242ICV500	ICVPO110425AR1242	PO114900.D	04 Nov 2025 20:21		YP/AJ	Ok,M
33	AR1248ICV500	ICVPO110425AR1248	PO114901.D	04 Nov 2025 20:58		YP/AJ	Ok
34	AR1254ICV500	ICVPO110425AR1254	PO114902.D	04 Nov 2025 21:35		YP/AJ	Ok
35	AR1268ICV500	ICVPO110425AR1268	PO114903.D	04 Nov 2025 22:12		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114904.D	04 Nov 2025 22:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO112125

Review By	rahul	Review On	11/21/2025 11:49:41 AM
Supervise By	mohammad	Supervise On	11/27/2025 12:19:53 AM
SubDirectory	PO112125	HP Acquire Method	HP Processing Method PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO115308.D	21 Nov 2025 08:38		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO115309.D	21 Nov 2025 08:57		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO115310.D	21 Nov 2025 09:22		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO115311.D	21 Nov 2025 09:40		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO115312.D	21 Nov 2025 09:59		YP/AJ	Ok,M
6	AR1268CCC500	AR1268CCC500	PO115313.D	21 Nov 2025 10:17		YP/AJ	Ok
7	I.BLK	I.BLK	PO115314.D	21 Nov 2025 10:35		YP/AJ	Ok
8	Q3666-02RE	BUR-25-0102	PO115315.D	21 Nov 2025 11:10	DCB high 1st column,TCMX high 2nd column	YP/AJ	Confirms
9	Q3673-05DL	C0AH0DL	PO115316.D	21 Nov 2025 11:28	AR1254+1260+1248 Hit	YP/AJ	Ok,M
10	Q3673-06DL	C0AH1DL	PO115317.D	21 Nov 2025 11:47	AR1254+1260 Hit	YP/AJ	Ok,M
11	Q3673-07DL2	C0AH2DL2	PO115318.D	21 Nov 2025 12:05	R1242+1260 Hit	YP/AJ	Ok,M
12	Q3697-01	MOO-25-0331	PO115319.D	21 Nov 2025 12:23	AR1260 Hit	YP/AJ	Ok,M
13	Q3697-02	MOO-25-0332	PO115320.D	21 Nov 2025 12:42	DCB high 2nd column	YP/AJ	Ok
14	Q3697-03	MOO-25-0333	PO115321.D	21 Nov 2025 13:00	DCB high 2nd column	YP/AJ	Ok
15	Q3699-01	TR-04-112025	PO115322.D	21 Nov 2025 13:19	AR1260 Hit	YP/AJ	Ok
16	Q3699-03	TR-06-112025	PO115323.D	21 Nov 2025 13:37	AR1260 Hit	YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO115324.D	21 Nov 2025 14:32		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO112125

Review By	rahul	Review On	11/21/2025 11:49:41 AM
Supervise By	mohammad	Supervise On	11/27/2025 12:19:53 AM
SubDirectory	PO112125	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO115325.D	21 Nov 2025 14:51		YP/AJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PO115326.D	21 Nov 2025 15:09		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO115327.D	21 Nov 2025 15:27		YP/AJ	Ok,M
21	AR1268CCC500	AR1268CCC500	PO115328.D	21 Nov 2025 15:46		YP/AJ	Ok
22	I.BLK	I.BLK	PO115329.D	21 Nov 2025 16:04		YP/AJ	Ok
23	PB170688BL	PB170688BL	PO115330.D	21 Nov 2025 16:22		YP/AJ	Ok
24	PB170688BS	PB170688BS	PO115331.D	21 Nov 2025 16:41		YP/AJ	Ok,M
25	Q3693-11	C0AL4	PO115332.D	21 Nov 2025 16:59	AR1260 Hit	YP/AJ	Ok,M
26	Q3693-12	C0AL5	PO115333.D	21 Nov 2025 17:18	AR1260 Hit	YP/AJ	Ok,M
27	Q3693-13	C0AL6	PO115334.D	21 Nov 2025 17:36	AR1260 Hit	YP/AJ	Ok
28	Q3693-14	C0AL7	PO115335.D	21 Nov 2025 17:54	AR1254 Hit	YP/AJ	Ok
29	Q3694-01	C0AD8	PO115336.D	21 Nov 2025 18:13	AR1260 Hit,Need 10X,100X	YP/AJ	Dilution
30	Q3694-01MS	C0AD8MS	PO115337.D	21 Nov 2025 18:31	AR1260 Recovery fail	YP/AJ	Not Ok
31	Q3694-01MSD	C0AD8MSD	PO115338.D	21 Nov 2025 18:50	AR1260 Recovery fail and conc high	YP/AJ	Not Ok
32	Q3694-02	C0AE1	PO115339.D	21 Nov 2025 19:08	AR1260 Hit, Need 2X	YP/AJ	Dilution
33	AR1660CCC500	AR1660CCC500	PO115340.D	21 Nov 2025 20:22		YP/AJ	Ok
34	AR1242CCC500	AR1242CCC500	PO115341.D	21 Nov 2025 21:17		YP/AJ	Ok,M
35	AR1248CCC500	AR1248CCC500	PO115342.D	21 Nov 2025 21:35		YP/AJ	Ok,M
36	AR1254CCC500	AR1254CCC500	PO115343.D	21 Nov 2025 21:53	AR1254-03 & DCB high in 1st column	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO112125

Review By	rahul	Review On	11/21/2025 11:49:41 AM
Supervise By	mohammad	Supervise On	11/27/2025 12:19:53 AM
SubDirectory	PO112125	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	I.BLK	I.BLK	PO115344.D	21 Nov 2025 22:12		YP/AJ	Ok
38	Q3694-03	C0AE2	PO115345.D	21 Nov 2025 22:30	AR1260 Hit	YP/AJ	Ok
39	Q3694-04	C0AE3	PO115346.D	21 Nov 2025 22:49	AR1242+1260 Hit	YP/AJ	Ok,M
40	Q3694-05	C0AF3	PO115347.D	21 Nov 2025 23:07	AR1260 Hit	YP/AJ	Ok
41	Q3694-06	C0AF4	PO115348.D	21 Nov 2025 23:26	AR1260 Hit	YP/AJ	Ok
42	Q3694-07	C0AF5	PO115349.D	21 Nov 2025 23:44	AR1242+AR1254+AR1260 hit	YP/AJ	Ok,M
43	Q3694-08	C0AF6	PO115350.D	22 Nov 2025 00:02	AR1242+AR1260 hit	YP/AJ	Ok,M
44	Q3694-09	C0AL8	PO115351.D	22 Nov 2025 00:21	AR1260 Hit	YP/AJ	Ok,M
45	Q3694-10	C0AL9	PO115352.D	22 Nov 2025 00:39	AR1254+1260 Hit ,	YP/AJ	Ok,M
46	Q3694-11	C0AM0	PO115353.D	22 Nov 2025 00:58	AR1260 Hit	YP/AJ	Ok,M
47	Q3694-12	C0AM1	PO115354.D	22 Nov 2025 01:16	AR1260 Hit,Need 5X	YP/AJ	Dilution
48	AR1660CCC500	AR1660CCC500	PO115355.D	22 Nov 2025 02:30		YP/AJ	Ok,M
49	AR1242CCC500	AR1242CCC500	PO115356.D	22 Nov 2025 03:25		YP/AJ	Ok,M
50	AR1248CCC500	AR1248CCC500	PO115357.D	22 Nov 2025 03:43		YP/AJ	Ok
51	AR1254CCC500	AR1254CCC500	PO115358.D	22 Nov 2025 04:02	dcb & tcm high in both column	YP/AJ	Not Ok
52	I.BLK	I.BLK	PO115359.D	22 Nov 2025 04:20		YP/AJ	Ok
53	Q3694-13	C0AM2	PO115360.D	22 Nov 2025 04:39	AR1260 Hit	YP/AJ	Ok
54	Q3694-14	C0AM3	PO115361.D	22 Nov 2025 04:57	AR1260 Hit	YP/AJ	Ok,M
55	Q3694-15	C0AM4	PO115362.D	22 Nov 2025 05:15	AR1260 Hit	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO112125

Review By	rahul	Review On	11/21/2025 11:49:41 AM
Supervise By	mohammad	Supervise On	11/27/2025 12:19:53 AM
SubDirectory	PO112125	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

56	Q3694-16	C0AM5	PO115363.D	22 Nov 2025 05:34	AR1260 Hit	YP/AJ	Ok
57	Q3673-07DL	C0AH2DL	PO115364.D	22 Nov 2025 05:52		YP/AJ	Not Ok
58	AR1660CCC500	AR1660CCC500	PO115365.D	22 Nov 2025 07:06		YP/AJ	Ok,M
59	AR1242CCC500	AR1242CCC500	PO115366.D	22 Nov 2025 08:01		YP/AJ	Ok,M
60	AR1248CCC500	AR1248CCC500	PO115367.D	22 Nov 2025 08:20		YP/AJ	Ok
61	AR1254CCC500	AR1254CCC500	PO115368.D	22 Nov 2025 08:38	dcb & tcm high in both column	YP/AJ	Not Ok
62	I.BLK	I.BLK	PO115369.D	22 Nov 2025 08:56		YP/AJ	Ok
63	DDT ANALOG	DDT ANALOG	PO115370.D	22 Nov 2025 09:15		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ111825

Review By	rahul	Review On	11/19/2025 8:33:53 AM
Supervise By	Sohil	Supervise On	11/19/2025 3:39:22 PM
SubDirectory	PQ111825	HP Acquire Method	HP Processing Method PQ111825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071271.D	18 Nov 2025 11:44		YP\AJ	Ok
2	I.BLK	I.BLK	PQ071272.D	18 Nov 2025 11:59		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ071273.D	18 Nov 2025 12:13		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ071274.D	18 Nov 2025 12:28		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ071275.D	18 Nov 2025 12:43		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ071276.D	18 Nov 2025 12:57		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ071277.D	18 Nov 2025 15:42		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ071278.D	18 Nov 2025 16:25		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ071279.D	18 Nov 2025 16:39		YP\AJ	Ok
10	AR1242ICC500	AR1242ICC500	PQ071280.D	18 Nov 2025 16:54		YP\AJ	Ok
11	AR1248ICC500	AR1248ICC500	PQ071281.D	18 Nov 2025 17:08		YP\AJ	Ok
12	AR1254ICC500	AR1254ICC500	PQ071282.D	18 Nov 2025 17:23		YP\AJ	Ok
13	AR1262ICC500	AR1262ICC500	PQ071283.D	18 Nov 2025 17:37		YP\AJ	Ok
14	AR1268ICC500	AR1268ICC500	PQ071284.D	18 Nov 2025 17:52		YP\AJ	Ok
15	PQ111825ICV500	ICVPQ111825	PQ071285.D	18 Nov 2025 18:07		YP\AJ	Ok
16	AR1660CCC500	AR1660CCC500	PQ071286.D	18 Nov 2025 18:36		YP\AJ	Ok
17	I.BLK	I.BLK	PQ071287.D	18 Nov 2025 19:05		YP\AJ	Ok
18	PB170532BL	PB170532BL	PQ071288.D	18 Nov 2025 19:20	DCB high in 2nd column	YP\AJ	Not Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ111825

Review By	rahul	Review On	11/19/2025 8:33:53 AM	
Supervise By	Sohil	Supervise On	11/19/2025 3:39:22 PM	
SubDirectory	PQ111825	HP Acquire Method	HP Processing Method	PQ111825
STD. NAME	STD REF.#			
Tune/Reschk				
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854			
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852			
Internal Standard/PEM				
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862			
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

19	Q3530-03	MDL-SOIL-03-QT4-202	PQ071289.D	18 Nov 2025 19:34	AR1660 MDL SOIL 25PPB	YPIAJ	Ok
20	Q3530-03	MDL-SOIL-03-QT4-202	PQ071290.D	18 Nov 2025 19:49	AR1221 MDL SOIL 25PPB, failing high	YPIAJ	Not Ok
21	Q3530-03	MDL-SOIL-03-QT4-202	PQ071291.D	18 Nov 2025 20:03	AR1232 MDL SOIL 25PPB, failing high in 1st column	YPIAJ	Not Ok
22	Q3530-03	MDL-SOIL-03-QT4-202	PQ071292.D	18 Nov 2025 20:18	AR1242 MDL SOIL 25PPB, failing high in 2d column	YPIAJ	Not Ok
23	Q3530-03	MDL-SOIL-03-QT4-202	PQ071293.D	18 Nov 2025 20:33	AR1248 MDL SOIL 25PPB, failing high	YPIAJ	Not Ok
24	Q3530-03	MDL-SOIL-03-QT4-202	PQ071294.D	18 Nov 2025 20:47	AR1254 MDL SOIL 25PPB, failing high	YPIAJ	Not Ok
25	Q3530-03	MDL-SOIL-03-QT4-202	PQ071295.D	18 Nov 2025 21:02	AR1262 MDL SOIL 25PPB	YPIAJ	Not Ok
26	Q3530-03	MDL-SOIL-03-QT4-202	PQ071296.D	18 Nov 2025 21:16	AR1268 MDL SOIL 25PPB	YPIAJ	Ok
27	AR1660CCC500	AR1660CCC500	PQ071297.D	18 Nov 2025 22:14		YPIAJ	Ok
28	I.BLK	I.BLK	PQ071298.D	19 Nov 2025 00:49		YPIAJ	Ok
29	PB170530BL	PB170530BL	PQ071299.D	19 Nov 2025 01:03		YPIAJ	Ok
30	Q3530-09	MDL-WATER-03-QT4-2	PQ071300.D	19 Nov 2025 01:18	AR1660 MDL WATER 25PPB	YPIAJ	Ok
31	Q3530-09	MDL-WATER-03-QT4-2	PQ071301.D	19 Nov 2025 01:32	AR1221 MDL WATER 25PPB, failing high	YPIAJ	Not Ok
32	Q3530-09	MDL-WATER-03-QT4-2	PQ071302.D	19 Nov 2025 01:47	AR1232 MDL WATER 25PPB, failing high in 1st column	YPIAJ	Not Ok
33	Q3530-09	MDL-WATER-03-QT4-2	PQ071303.D	19 Nov 2025 02:01	AR1242 MDL WATER 25PPB, failing high in 2d column	YPIAJ	Not Ok
34	Q3530-09	MDL-WATER-03-QT4-2	PQ071304.D	19 Nov 2025 02:16	AR1248 MDL WATER 25PPB, failing high	YPIAJ	Not Ok
35	Q3530-09	MDL-WATER-03-QT4-2	PQ071305.D	19 Nov 2025 02:31	AR1254 MDL WATER 25PPB, failing high	YPIAJ	Not Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ111825

Review By	rahul	Review On	11/19/2025 8:33:53 AM	
Supervise By	Sohil	Supervise On	11/19/2025 3:39:22 PM	
SubDirectory	PQ111825	HP Acquire Method	HP Processing Method	PQ111825
STD. NAME	STD REF.#			
Tune/Reschk	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862			
Initial Calibration Stds				
CCC				
Internal Standard/PEM				
ICV/I.BLK				
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

36	Q3530-09	MDL-WATER-03-QT4-2	PQ071306.D	19 Nov 2025 02:45	AR1262 MDL WATER 25PPB	YPIAJ	Ok,M
37	Q3530-09	MDL-WATER-03-QT4-2	PQ071307.D	19 Nov 2025 03:00	AR1268 MDL WATER 25PPB	YPIAJ	Ok
38	AR1660CCC500	AR1660CCC500	PQ071308.D	19 Nov 2025 03:58		YPIAJ	Ok
39	I.BLK	I.BLK	PQ071309.D	19 Nov 2025 04:30		YPIAJ	Ok
40	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071310.D	19 Nov 2025 04:44	AR1660 LOQ SOIL 50PPB, failing low	YPIAJ	Not Ok
41	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071311.D	19 Nov 2025 04:59	AR1221 LOQ SOIL 50PPB	YPIAJ	Ok,M
42	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071312.D	19 Nov 2025 05:14	AR1232 LOQ SOIL 50PPB	YPIAJ	Ok,M
43	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071313.D	19 Nov 2025 05:28	AR1242 LOQ SOIL 50PPB	YPIAJ	Ok,M
44	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071314.D	19 Nov 2025 05:43	AR1248 LOQ SOIL 50PPB	YPIAJ	Ok
45	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071315.D	19 Nov 2025 05:57	AR1254 LOQ SOIL 50PPB	YPIAJ	Ok,M
46	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071316.D	19 Nov 2025 06:12	AR1262 LOQ SOIL 50PPB	YPIAJ	Ok
47	Q3530-02	LOQ-SOIL-02-QT4-202	PQ071317.D	19 Nov 2025 06:27	AR1268 LOQ SOIL 50PPB	YPIAJ	Ok,M
48	AR1660CCC500	AR1660CCC500	PQ071318.D	19 Nov 2025 07:25		YPIAJ	Ok
49	I.BLK	I.BLK	PQ071319.D	19 Nov 2025 08:09		YPIAJ	Ok
50	Q3530-08	LOQ-WATER-02-QT4-2	PQ071320.D	19 Nov 2025 08:23	AR1660 LOQ WATER 50PPB , Recovery low	YPIAJ	Not Ok
51	Q3530-08	LOQ-WATER-02-QT4-2	PQ071321.D	19 Nov 2025 08:38	AR1221 LOQ WATER 50PPB	YPIAJ	Ok,M
52	Q3530-08	LOQ-WATER-02-QT4-2	PQ071322.D	19 Nov 2025 08:52	AR1232 LOQ WATER 50PPB	YPIAJ	Ok,M
53	Q3530-08	LOQ-WATER-02-QT4-2	PQ071323.D	19 Nov 2025 09:07	AR1242 LOQ WATER 50PPB	YPIAJ	Ok,M
54	Q3530-08	LOQ-WATER-02-QT4-2	PQ071324.D	19 Nov 2025 09:21	AR1248 LOQ WATER 50PPB	YPIAJ	Ok,M

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ111825

Review By	rahul	Review On	11/19/2025 8:33:53 AM
Supervise By	Sohil	Supervise On	11/19/2025 3:39:22 PM
SubDirectory	PQ111825	HP Acquire Method	HP Processing Method PQ111825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

55	Q3530-08	LOQ-WATER-02-QT4-2	PQ071325.D	19 Nov 2025 09:36	AR1254 LOQ WATER 50PPB , AR1254-1 high in 1st column	YPIAJ	Not Ok
56	Q3530-08	LOQ-WATER-02-QT4-2	PQ071326.D	19 Nov 2025 09:51	AR1262 LOQ WATER 50PPB	YPIAJ	Ok,M
57	Q3530-08	LOQ-WATER-02-QT4-2	PQ071327.D	19 Nov 2025 10:05	AR1268 LOQ WATER 50PPB , AR1268-1,2 High in 2nd column , all surrogate fail	YPIAJ	Not Ok
58	AR1660CCC500	AR1660CCC500	PQ071328.D	19 Nov 2025 11:03		YPIAJ	Ok
59	I.BLK	I.BLK	PQ071329.D	19 Nov 2025 11:19		YPIAJ	Ok
60	DDT ANALOG	DDT ANALOG	PQ071330.D	19 Nov 2025 11:33		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ112125

Review By	rahul	Review On	11/21/2025 11:49:09 AM
Supervise By	mohammad	Supervise On	11/26/2025 4:59:58 AM
SubDirectory	PQ112125	HP Acquire Method	HP Processing Method PQ111825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071417.D	21 Nov 2025 08:26		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071418.D	21 Nov 2025 09:06		YP\AJ	Ok,M
3	I.BLK	I.BLK	PQ071419.D	21 Nov 2025 10:09		YP\AJ	Ok
4	Q3672-02DL	C0AE5DL	PQ071420.D	21 Nov 2025 10:27	AR1260+1248 Hit	YP\AJ	Ok,M
5	Q3672-03DL	C0AE6DL	PQ071421.D	21 Nov 2025 10:42	AR1260 Hit	YP\AJ	Ok
6	Q3672-04DL	C0AE7DL	PQ071422.D	21 Nov 2025 10:57	AR1260, Need more dilution	YP\AJ	Dilution
7	Q3672-05DL	C0AE8DL	PQ071423.D	21 Nov 2025 11:11	AR1260 Hit	YP\AJ	Ok
8	Q3672-11DL	C0AF8DL	PQ071424.D	21 Nov 2025 11:26	AR1260+1254+1248 Hit	YP\AJ	Ok
9	Q3672-12DL	C0AF9DL	PQ071425.D	21 Nov 2025 11:40	AR1248+AR1260 Hit,Need more dilution	YP\AJ	Dilution
10	Q3672-13DL	C0AG0DL	PQ071426.D	21 Nov 2025 11:55	AR1260 hit,Need more dilution	YP\AJ	Not Ok
11	Q3674-01	C0AH9	PQ071427.D	21 Nov 2025 12:16	AR1260 Hit	YP\AJ	Ok
12	Q3674-02	C0AJ0	PQ071428.D	21 Nov 2025 12:31	AR1260 Hit,Need 5X	YP\AJ	Dilution
13	Q3674-03	C0AJ1	PQ071429.D	21 Nov 2025 12:45	AR1260 Hit,Need 50X	YP\AJ	Dilution
14	AR1660CCC500	AR1660CCC500	PQ071430.D	21 Nov 2025 13:43		YP\AJ	Ok,M
15	I.BLK	I.BLK	PQ071431.D	21 Nov 2025 13:59		YP\AJ	Ok
16	PB170671BL	PB170671BL	PQ071432.D	21 Nov 2025 14:13		YP\AJ	Ok
17	PB170671BS	PB170671BS	PQ071433.D	21 Nov 2025 14:28		YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ112125

Review By	rahul	Review On	11/21/2025 11:49:09 AM
Supervise By	mohammad	Supervise On	11/26/2025 4:59:58 AM
SubDirectory	PQ112125	HP Acquire Method	HP Processing Method PQ111825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q3674-04	C0AJ2	PQ071434.D	21 Nov 2025 14:42	AR1260 Hit,Need 10X	YP\AJ	Dilution
19	Q3674-05	C0AJ3	PQ071435.D	21 Nov 2025 14:57	AR1254+AR1260 Hit	YP\AJ	Ok
20	Q3674-06	C0AJ4	PQ071436.D	21 Nov 2025 15:11	AR1254+AR1260 Hit	YP\AJ	Ok
21	Q3674-07	C0AJ5	PQ071437.D	21 Nov 2025 15:26	AR248+AR1254+AR1260 Hit, Need 2x	YP\AJ	Dilution
22	Q3674-08	C0AJ6	PQ071438.D	21 Nov 2025 15:41	AR1254+AR1260 Hit	YP\AJ	Ok
23	Q3674-09	C0AJ7	PQ071439.D	21 Nov 2025 15:55	AR1260 Hit	YP\AJ	Ok,M
24	Q3674-09MS	C0AJ7MS	PQ071440.D	21 Nov 2025 16:10		YP\AJ	Ok
25	Q3674-09MSD	C0AJ7MSD	PQ071441.D	21 Nov 2025 16:24		YP\AJ	Ok
26	AR1660CCC500	AR1660CCC500	PQ071442.D	21 Nov 2025 17:23		YP\AJ	Ok,M
27	I.BLK	I.BLK	PQ071443.D	21 Nov 2025 18:06		YP\AJ	Ok
28	Q3674-10	C0AJ8	PQ071444.D	21 Nov 2025 18:21	AR1260 Hit	YP\AJ	Ok
29	Q3674-11	C0AJ9	PQ071445.D	21 Nov 2025 18:36	AR1260 Hit	YP\AJ	Ok,M
30	Q3674-12	C0AK0	PQ071446.D	21 Nov 2025 18:50	AR1260 Hit	YP\AJ	Ok
31	Q3674-13	C0AK1	PQ071447.D	21 Nov 2025 19:05	AR1254 Hit	YP\AJ	Ok,M
32	Q3674-14	C0AK3	PQ071448.D	21 Nov 2025 19:19	AR1260 Hit, Need 50x	YP\AJ	Dilution
33	Q3674-15	C0AK4	PQ071449.D	21 Nov 2025 19:34	AR1260 Hit	YP\AJ	Ok
34	Q3693-01	C0AJ7	PQ071450.D	21 Nov 2025 19:48	AR1260 Hit	YP\AJ	Ok
35	Q3693-01MS	C0AJ7MS	PQ071451.D	21 Nov 2025 20:03		YP\AJ	Ok
36	Q3693-01MSD	C0AJ7MSD	PQ071452.D	21 Nov 2025 20:18	RPD fail for AR1260	YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ112125

Review By	rahul	Review On	11/21/2025 11:49:09 AM
Supervise By	mohammad	Supervise On	11/26/2025 4:59:58 AM
SubDirectory	PQ112125	HP Acquire Method	HP Processing Method PQ111825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q3693-02	C0AK5	PQ071453.D	21 Nov 2025 20:32	AR1260 Hit	YPIAJ	Ok
38	AR1660CCC500	AR1660CCC500	PQ071454.D	21 Nov 2025 21:31		YPIAJ	Ok,M
39	I.BLK	I.BLK	PQ071455.D	21 Nov 2025 22:29		YPIAJ	Ok
40	PB170682BL	PB170682BL	PQ071456.D	21 Nov 2025 22:44		YPIAJ	Ok
41	PB170682BS	PB170682BS	PQ071457.D	21 Nov 2025 22:58		YPIAJ	Ok
42	Q3693-03	C0AK6	PQ071458.D	21 Nov 2025 23:13	AR1260 hits,Need 50x,500X,DCB high both column	YPIAJ	Dilution
43	Q3693-04	C0AK7	PQ071459.D	21 Nov 2025 23:27	AR1260 Hit, Need 5X	YPIAJ	Dilution
44	Q3693-05	C0AK8	PQ071460.D	21 Nov 2025 23:42	AR1260 Hit	YPIAJ	Ok
45	Q3693-06	C0AK9	PQ071461.D	21 Nov 2025 23:57	AR248+AR1254+AR1260 Hit	YPIAJ	Ok
46	Q3693-07	C0AL0	PQ071462.D	22 Nov 2025 00:11	AR248+AR1254+AR1260 Hit	YPIAJ	Ok
47	Q3693-08	C0AL1	PQ071463.D	22 Nov 2025 00:26	AR248+AR1254+AR1260 Hit	YPIAJ	Ok
48	Q3693-09	C0AL2	PQ071464.D	22 Nov 2025 00:40	AR248+AR1254+AR1260 Hit	YPIAJ	Ok
49	Q3693-10	C0AL3	PQ071465.D	22 Nov 2025 00:55	AR1260 Hit	YPIAJ	Ok
50	AR1660CCC500	AR1660CCC500	PQ071466.D	22 Nov 2025 01:53		YPIAJ	Ok,M
51	I.BLK	I.BLK	PQ071467.D	22 Nov 2025 02:52		YPIAJ	Ok
52	DDT ANALOG	DDT ANALOG	PQ071468.D	22 Nov 2025 03:06		YPIAJ	Ok
53	Q3672-04DL2	C0AE7DL2	PQ071469.D	22 Nov 2025 03:21	AR1260 Hit	YPIAJ	Ok
54	Q3672-13DL2	C0AG0DL2	PQ071470.D	22 Nov 2025 03:36	AR1260 hit,Need more dilution	YPIAJ	Not Ok
55	AR1660CCC500	AR1660CCC500	PQ071471.D	22 Nov 2025 04:34		YPIAJ	Ok,M

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ112125

Review By	rahul	Review On	11/21/2025 11:49:09 AM				
Supervise By	mohammad	Supervise On	11/26/2025 4:59:58 AM				
SubDirectory	PQ112125	HP Acquire Method	HP Processing Method		PQ111825		
STD. NAME		STD REF.#					
Tune/Reschk		PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862					
Initial Calibration Stds							
CCC							
Internal Standard/PEM							
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							
56	I.BLK	I.BLK	PQ071472.D	22 Nov 2025 05:32		YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ112525

Review By	rahul	Review On	11/25/2025 2:24:11 PM
Supervise By	mohammad	Supervise On	11/27/2025 12:20:41 AM
SubDirectory	PQ112525	HP Acquire Method	HP Processing Method PQ111825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071498.D	25 Nov 2025 09:00		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071499.D	25 Nov 2025 09:15		YP\AJ	Ok
3	I.BLK	I.BLK	PQ071500.D	25 Nov 2025 09:30		YP\AJ	Ok
4	Q3693-03DL	C0AK6DL	PQ071501.D	25 Nov 2025 10:56	AR1260 hits,Need more dilution	YP\AJ	Dilution
5	Q3693-03DL2	C0AK6DL2	PQ071502.D	25 Nov 2025 11:11	AR1260 hits	YP\AJ	Ok
6	Q3693-04DL	C0AK7DL	PQ071503.D	25 Nov 2025 11:25	AR1260 hits	YP\AJ	Ok
7	Q3679-01	111425-A	PQ071504.D	25 Nov 2025 11:53		YP\AJ	Ok
8	Q3679-02	111825	PQ071505.D	25 Nov 2025 12:08	AR1242 Hit , AR1242 CCAL missing , need 5x dilution	YP\AJ	Not Ok
9	Q3713-01	1	PQ071506.D	25 Nov 2025 12:23	AR1254 Hit , AR1254 CCAL missing	YP\AJ	Not Ok
10	Q3713-02	2	PQ071507.D	25 Nov 2025 12:37	AR1254 Hit , AR1254 CCAL missing	YP\AJ	Not Ok
11	Q3713-03	3	PQ071508.D	25 Nov 2025 12:52	AR1254 Hit , AR1254 CCAL missing	YP\AJ	Not Ok
12	Q3713-04	4	PQ071509.D	25 Nov 2025 13:06	AR1254 Hit , AR1254 CCAL missing	YP\AJ	Not Ok
13	Q3713-05	5	PQ071510.D	25 Nov 2025 13:21	AR1254 Hit , AR1254 CCAL missing	YP\AJ	Not Ok
14	AR1660CCC500	AR1660CCC500	PQ071511.D	25 Nov 2025 13:50		YP\AJ	Ok
15	I.BLK	I.BLK	PQ071512.D	25 Nov 2025 14:05		YP\AJ	Ok
16	PB170728BL	PB170728BL	PQ071513.D	25 Nov 2025 14:19		YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ112525

Review By	rahul	Review On	11/25/2025 2:24:11 PM
Supervise By	mohammad	Supervise On	11/27/2025 12:20:41 AM
SubDirectory	PQ112525	HP Acquire Method	HP Processing Method PQ111825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	PB170728BS	PB170728BS	PQ071514.D	25 Nov 2025 14:34		YPIAJ	Ok
18	Q3717-01	1A1B1C	PQ071515.D	25 Nov 2025 14:48		YPIAJ	Ok,M
19	Q3717-02	2A2B2C	PQ071516.D	25 Nov 2025 15:03		YPIAJ	Ok,M
20	Q3717-03	3A3B3C	PQ071517.D	25 Nov 2025 15:18		YPIAJ	Ok,M
21	Q3717-04	4A4B4C	PQ071518.D	25 Nov 2025 15:32		YPIAJ	Ok,M
22	Q3717-05	5A5B5C	PQ071519.D	25 Nov 2025 15:47		YPIAJ	Ok,M
23	Q3717-06	6A6B6C	PQ071520.D	25 Nov 2025 16:01		YPIAJ	Ok
24	Q3717-07	7A7B7C	PQ071521.D	25 Nov 2025 16:16		YPIAJ	Ok,M
25	Q3717-08	8A8B8C	PQ071522.D	25 Nov 2025 16:30		YPIAJ	Ok
26	AR1660CCC500	AR1660CCC500	PQ071523.D	25 Nov 2025 17:00		YPIAJ	Ok
27	I.BLK	I.BLK	PQ071524.D	25 Nov 2025 17:14		YPIAJ	Ok
28	PB170731BL	PB170731BL	PQ071525.D	25 Nov 2025 17:29		YPIAJ	Ok
29	PB170731BS	PB170731BS	PQ071526.D	25 Nov 2025 17:43		YPIAJ	Ok
30	PB170731BSD	PB170731BSD	PQ071527.D	25 Nov 2025 17:58		YPIAJ	Ok
31	Q3697-04	MOO-25-0329	PQ071528.D	25 Nov 2025 18:13		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PQ071529.D	25 Nov 2025 18:42		YPIAJ	Ok
33	I.BLK	I.BLK	PQ071530.D	25 Nov 2025 19:25		YPIAJ	Ok

M : Manual Integration