

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170688BS

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: PB170688BS

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115331.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.479	5.429	5.529	149	149	7.67
COLUMN 1	2	5.501	5.451	5.551	148		
	3	5.562	5.512	5.612	154		
	4	5.658	5.608	5.708	152		
	5	5.95	5.9	6	142		
	1	4.72	4.67	4.77	139		
COLUMN 2	2	4.739	4.689	4.789	140	138	
	3	4.913	4.863	4.963	139		
	4	4.956	4.906	5.006	137		
	5	5.167	5.117	5.217	137		
Aroclor-1260	1	7.063	7.013	7.113	142	135	7.69
COLUMN 1	2	7.316	7.266	7.366	149		
	3	7.674	7.624	7.724	127		
	4	7.896	7.846	7.946	131		
	5	8.202	8.152	8.252	127		
	1	6.197	6.147	6.247	134		
COLUMN 2	2	6.385	6.335	6.435	126	125	
	3	6.537	6.487	6.587	132		
	4	7.008	6.958	7.058	115		
	5	7.249	7.199	7.299	116		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AD8

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-01

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115336.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.064	7.014	7.114	20300	21000	10
COLUMN 1	2	7.322	7.272	7.372	23900		
	3	7.68	7.63	7.73	21800		
	4	7.893	7.843	7.943	23200		
	5	8.196	8.146	8.246	18000		
	1	6.198	6.148	6.248	19900		
COLUMN 2	2	6.386	6.336	6.436	19200	19000	
	3	6.538	6.488	6.588	20400		
	4	7.009	6.959	7.059	20200		
	5	7.253	7.203	7.303	17500		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AE1

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-02

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115339.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.062	7.012	7.112	533	539	13.47
COLUMN 1	2	7.317	7.267	7.367	575		
	3	7.674	7.624	7.724	534		
	4	7.896	7.846	7.946	524		
	5	8.202	8.152	8.252	527		
	1	6.197	6.147	6.247	484		
COLUMN 2	2	6.384	6.334	6.434	470	471	
	3	6.537	6.487	6.587	471		
	4	7.007	6.957	7.057	457		
	5	7.249	7.199	7.299	472		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AE2

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-03

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115345.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	127	131	11.29
COLUMN 1	2	7.317	7.267	7.367	137		
	3	7.674	7.624	7.724	127		
	4	7.896	7.846	7.946	133		
	5	8.203	8.153	8.253	129		
	1	6.197	6.147	6.247	120		
COLUMN 2	2	6.385	6.335	6.435	115	117	
	3	6.537	6.487	6.587	117		
	4	7.008	6.958	7.058	117		
	5	7.249	7.199	7.299	117		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AE3

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-04

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115346.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.479	5.429	5.529	41.7	48.7	7.7
COLUMN 1	2	5.501	5.451	5.551	67.8		
	3	5.57	5.52	5.62	15.5		
	4	5.658	5.608	5.708	68.5		
	5	6.385	6.335	6.435	50.1		
	1	4.72	4.67	4.77	42.7		
COLUMN 2	2	4.739	4.689	4.789	62.6	52.6	
	3	4.913	4.863	4.963	59.4		
	4	4.997	4.947	5.047	52.6		
	5	5.517	5.467	5.567	45.7		
Aroclor-1260	1	7.063	7.013	7.113	10.7	11.1	11.43
COLUMN 1	2	7.317	7.267	7.367	10.6		
	3	7.673	7.623	7.723	9.70		
	4	7.893	7.843	7.943	13.1		
	5	8.202	8.152	8.252	11.4		
	1	6.197	6.147	6.247	14.6		
COLUMN 2	2	6.384	6.334	6.434	10.2	9.90	
	3	6.536	6.486	6.586	7.91		
	4	7.007	6.957	7.057	7.99		
	5	7.248	7.198	7.298	8.77		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AF3

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-05

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115347.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	60.5		
	2	7.318	7.268	7.368	71.7		
	3	7.675	7.625	7.725	72.3		
	4	7.896	7.846	7.946	67.1		
	5	8.203	8.153	8.253	57.9		
COLUMN 1						65.9	
	1	6.197	6.147	6.247	58.3		
	2	6.385	6.335	6.435	61.9		
	3	6.538	6.488	6.588	54.4		
	4	7.008	6.958	7.058	58.1		
5	7.249	7.199	7.299	52.2			
COLUMN 2						57.0	14.48

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AF4

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-06

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115348.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	42.8	47.6	11.79
COLUMN 1	2	7.316	7.266	7.366	52.6		
	3	7.674	7.624	7.724	51.2		
	4	7.896	7.846	7.946	48.5		
	5	8.203	8.153	8.253	43.0		
	1	6.197	6.147	6.247	44.1	42.3	
COLUMN 2	2	6.384	6.334	6.434	46.7		
	3	6.537	6.487	6.587	40.0		
	4	7.007	6.957	7.057	41.9		
	5	7.248	7.198	7.298	38.9		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AF5

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-07

Date(s) Analyzed: 11/21/2025 11/21/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115349.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.48	5.43	5.53	52.3	64.0	33.01
COLUMN 1	2	5.501	5.451	5.551	86.9		
	3	5.567	5.517	5.617	18.7		
	4	5.658	5.608	5.708	88.1		
	5	6.385	6.335	6.435	73.9		
COLUMN 2	1	4.72	4.67	4.77	48.8	89.3	
	2	4.739	4.689	4.789	85.4		
	3	4.914	4.864	4.964	82.9		
	4	4.997	4.947	5.047	92.3		
	5	5.518	5.468	5.568	137		
Aroclor-1254	1	6.322	6.272	6.372	74.0	90.2	12.11
COLUMN 1	2	6.537	6.487	6.587	70.4		
	3	6.899	6.849	6.949	67.8		
	4	7.181	7.131	7.231	71.8		
	5	7.595	7.545	7.645	167		
COLUMN 2	1	5.518	5.468	5.568	65.4	79.9	
	2	5.665	5.615	5.715	65.6		
	3	6.067	6.017	6.117	85.4		
	4	6.294	6.244	6.344	63.8		
	5	6.71	6.66	6.76	119		

IDENTIFICATION SUMMARY FOR MULTICOMPONENT ANALYTES

Aroclor-1260	1	7.064	7.014	7.114	83.1	82.5	11.95
	2	7.317	7.267	7.367	96.8		
	3	7.675	7.625	7.725	72.2		
	4	7.896	7.846	7.946	99.0		
	5	8.203	8.153	8.253	61.3		
COLUMN 1	1	6.197	6.147	6.247	89.4	73.2	
	2	6.385	6.335	6.435	80.1		
	3	6.536	6.486	6.586	75.6		
	4	7.008	6.958	7.058	57.6		
	5	7.249	7.199	7.299	63.4		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AF6

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-08

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115350.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.48	5.43	5.53	26.5	32.5	19.69
COLUMN 1	2	5.501	5.451	5.551	48.7		
	3	5.567	5.517	5.617	12.0		
	4	5.659	5.609	5.709	53.3		
	5	6.385	6.335	6.435	21.8		
	1	4.72	4.67	4.77	27.3		
COLUMN 2	2	4.739	4.689	4.789	47.9	39.6	
	3	4.914	4.864	4.964	45.2		
	4	4.997	4.947	5.047	42.3		
	5	5.518	5.468	5.568	35.2		
Aroclor-1260	1	7.064	7.014	7.114	31.7	33.1	9.49
COLUMN 1	2	7.318	7.268	7.368	34.2		
	3	7.675	7.625	7.725	30.2		
	4	7.895	7.845	7.945	37.8		
	5	8.203	8.153	8.253	31.7		
	1	6.197	6.147	6.247	33.0		
COLUMN 2	2	6.385	6.335	6.435	31.0	30.1	
	3	6.537	6.487	6.587	30.1		
	4	7.008	6.958	7.058	28.2		
	5	7.248	7.198	7.298	28.4		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

COAL8

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-09

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115351.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.064	7.014	7.114	40.2	46.7	16.2
COLUMN 1	2	7.318	7.268	7.368	47.8		
	3	7.674	7.624	7.724	49.8		
	4	7.895	7.845	7.945	53.6		
	5	8.203	8.153	8.253	41.8		
	1	6.197	6.147	6.247	41.5		
COLUMN 2	2	6.385	6.335	6.435	42.6	39.7	
	3	6.538	6.488	6.588	38.1		
	4	7.008	6.958	7.058	39.8		
	5	7.249	7.199	7.299	36.4		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AL9

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-10

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115352.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.322	6.272	6.372	222	251	16.38
COLUMN 1	2	6.538	6.488	6.588	221		
	3	6.899	6.849	6.949	198		
	4	7.18	7.13	7.23	195		
	5	7.595	7.545	7.645	419		
	1	5.518	5.468	5.568	166		
COLUMN 2	2	5.665	5.615	5.715	201	213	
	3	6.067	6.017	6.117	243		
	4	6.295	6.245	6.345	167		
	5	6.71	6.66	6.76	289		
Aroclor-1260	1	7.063	7.013	7.113	242	233	14.75
COLUMN 1	2	7.317	7.267	7.367	265		
	3	7.674	7.624	7.724	205		
	4	7.896	7.846	7.946	269		
	5	8.203	8.153	8.253	185		
	1	6.197	6.147	6.247	245		
COLUMN 2	2	6.384	6.334	6.434	209	201	
	3	6.537	6.487	6.587	209		
	4	7.008	6.958	7.058	166		
	5	7.249	7.199	7.299	178		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

COAM0

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-11

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115353.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	39.3	43.5	14.29
COLUMN 1	2	7.317	7.267	7.367	46.9		
	3	7.674	7.624	7.724	46.6		
	4	7.895	7.845	7.945	48.2		
	5	8.202	8.152	8.252	36.3		
	1	6.196	6.146	6.246	39.4		
COLUMN 2	2	6.384	6.334	6.434	41.2	37.7	
	3	6.537	6.487	6.587	34.7		
	4	7.007	6.957	7.057	38.1		
	5	7.249	7.199	7.299	35.2		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AM1

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-12

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115354.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.064	7.014	7.114	908	1100	21.22
COLUMN 1	2	7.317	7.267	7.367	1140		
	3	7.675	7.625	7.725	1130		
	4	7.896	7.846	7.946	1140		
	5	8.202	8.152	8.252	1110		
	1	6.198	6.148	6.248	806		
COLUMN 2	2	6.385	6.335	6.435	873	889	
	3	6.538	6.488	6.588	812		
	4	7.008	6.958	7.058	950		
	5	7.249	7.199	7.299	1000		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AM2

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-13

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115360.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	64.9	63.6	9.9
COLUMN 1	2	7.317	7.267	7.367	68.2		
	3	7.675	7.625	7.725	60.4		
	4	7.896	7.846	7.946	68.0		
	5	8.202	8.152	8.252	56.6		
	1	6.197	6.147	6.247	65.2		
COLUMN 2	2	6.385	6.335	6.435	59.0	57.6	
	3	6.537	6.487	6.587	58.3		
	4	7.008	6.958	7.058	54.6		
	5	7.249	7.199	7.299	50.9		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AM3

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-14

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115361.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	27.4	28.0	12.52
COLUMN 1	2	7.318	7.268	7.368	29.7		
	3	7.676	7.626	7.726	28.5		
	4	7.895	7.845	7.945	29.4		
	5	8.203	8.153	8.253	25.0		
	1	6.197	6.147	6.247	28.3		
COLUMN 2	2	6.385	6.335	6.435	27.0	24.7	
	3	6.537	6.487	6.587	24.5		
	4	7.008	6.958	7.058	21.5		
	5	7.249	7.199	7.299	22.1		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AM4

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-15

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115362.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.063	7.013	7.113	16.6	13.5	1.47
COLUMN 1	2	7.318	7.268	7.368	12.9		
	3	7.674	7.624	7.724	11.1		
	4	7.894	7.844	7.944	15.7		
	5	8.202	8.152	8.252	11.2		
	1	6.197	6.147	6.247	14.3		
COLUMN 2	2	6.385	6.335	6.435	25.8	13.7	
	3	6.537	6.487	6.587	10.7		
	4	7.008	6.958	7.058	9.15		
	5	7.249	7.199	7.299	8.47		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AM5

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-16

Date(s) Analyzed: 11/22/2025 11/22/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115363.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.064	7.014	7.114	11.1	9.30	11.36
COLUMN 1	2	7.318	7.268	7.368	8.75		
	3	7.675	7.625	7.725	6.60		
	4	7.896	7.846	7.946	12.3		
	5	8.204	8.154	8.254	7.65		
	1	6.197	6.147	6.247	13.4		
COLUMN 2	2	6.384	6.334	6.434	8.77	8.30	
	3	6.537	6.487	6.587	7.60		
	4	7.007	6.957	7.057	6.57		
	5	7.248	7.198	7.298	4.96		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AD8DL

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-01DL

Date(s) Analyzed: 11/25/2025 11/25/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115400.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.06	7.01	7.11	19100	20000	22.22
COLUMN 1	2	7.315	7.265	7.365	21400		
	3	7.671	7.621	7.721	19900		
	4	7.892	7.842	7.942	19300		
	5	8.199	8.149	8.249	21100		
	1	6.194	6.144	6.244	16500		
COLUMN 2	2	6.382	6.332	6.432	15900	16000	
	3	6.534	6.484	6.584	16400		
	4	7.005	6.955	7.055	15900		
	5	7.245	7.195	7.295	17200		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AD8DL2

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-01DL2

Date(s) Analyzed: 11/25/2025 11/25/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115401.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.061	7.011	7.111	18300	18000	11.76
COLUMN 1	2	7.314	7.264	7.364	19600		
	3	7.672	7.622	7.722	18000		
	4	7.894	7.844	7.944	17600		
	5	8.2	8.15	8.25	17800		
	1	6.194	6.144	6.244	16800	16000	
COLUMN 2	2	6.382	6.332	6.432	15800		
	3	6.534	6.484	6.584	16200		
	4	7.005	6.955	7.055	15300		
	5	7.246	7.196	7.296	15300		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AE1DL

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-02DL

Date(s) Analyzed: 11/25/2025 11/25/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115402.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.06	7.01	7.11	502	506	15.09
COLUMN 1	2	7.315	7.265	7.365	553		
	3	7.672	7.622	7.722	503		
	4	7.893	7.843	7.943	486		
	5	8.2	8.15	8.25	485		
	1	6.194	6.144	6.244	459		
COLUMN 2	2	6.382	6.332	6.432	437	435	
	3	6.535	6.485	6.585	438		
	4	7.005	6.955	7.055	419		
	5	7.246	7.196	7.296	425		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AM1DL

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-12DL

Date(s) Analyzed: 11/25/2025 11/25/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115410.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1260	1	7.06	7.01	7.11	872	1000	16.45
COLUMN 1	2	7.314	7.264	7.364	1080		
	3	7.672	7.622	7.722	1070		
	4	7.893	7.843	7.943	1100		
	5	8.199	8.149	8.249	1010		
	1	6.194	6.144	6.244	809		
COLUMN 2	2	6.382	6.332	6.432	846	848	
	3	6.534	6.484	6.584	783		
	4	7.005	6.955	7.055	896		
	5	7.246	7.196	7.296	904		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AD8MS

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-01MS

Date(s) Analyzed: 12/01/2025 12/01/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115470.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.479	5.429	5.529	210	219	6.19
COLUMN 1	2	5.5	5.45	5.55	220		
	3	5.561	5.511	5.611	207		
	4	5.657	5.607	5.707	224		
	5	5.948	5.898	5.998	232		
	1	4.717	4.667	4.767	201		
COLUMN 2	2	4.737	4.687	4.787	209	233	
	3	4.91	4.86	4.96	197		
	4	4.953	4.903	5.003	338		
	5	5.165	5.115	5.215	222		
	Aroclor-1260	1	7.062	7.012	7.112		4000
COLUMN 1	2	7.316	7.266	7.366	4430		
	3	7.672	7.622	7.722	3980		
	4	7.894	7.844	7.944	3800		
	5	8.201	8.151	8.251	4040		
	1	6.194	6.144	6.244	3290	3300	
COLUMN 2	2	6.381	6.331	6.431	3120		
	3	6.533	6.483	6.583	3320		
	4	7.004	6.954	7.054	3180		
	5	7.245	7.195	7.295	3390		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

C0AD8MSD

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Lab Sample ID: Q3694-01MSD

Date(s) Analyzed: 12/01/2025 12/01/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO115471.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.476	5.426	5.526	209	219	4.9
COLUMN 1	2	5.498	5.448	5.548	220		
	3	5.56	5.51	5.61	207		
	4	5.656	5.606	5.706	225		
	5	5.946	5.896	5.996	234		
	1	4.716	4.666	4.766	184		
COLUMN 2	2	4.736	4.686	4.786	210	230	
	3	4.91	4.86	4.96	196		
	4	4.952	4.902	5.002	337		
	5	5.163	5.113	5.213	223		
Aroclor-1260	1	7.06	7.01	7.11	3990	4100	21.62
COLUMN 1	2	7.313	7.263	7.363	4440		
	3	7.671	7.621	7.721	3990		
	4	7.893	7.843	7.943	3800		
	5	8.198	8.148	8.248	4040		
	1	6.192	6.142	6.242	3290		
COLUMN 2	2	6.38	6.33	6.43	3130	3300	
	3	6.533	6.483	6.583	3320		
	4	7.003	6.953	7.053	3180		
	5	7.245	7.195	7.295	3380		